

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073200.D
 Acq On : 20 May 2022 23:13
 Operator : VA/SY
 Sample : N2885-02
 Misc : 6.11G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P004-SS002-0612-01

Quant Time: May 21 00:18:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	78202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	130225	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	154075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	68641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	67058	77.592	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	155.180%
35) Dibromofluoromethane	7.908	113	56951	66.275	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	132.560%
50) Toluene-d8	10.332	98	163834	51.721	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.440%
62) 4-Bromofluorobenzene	12.620	95	66024	59.393	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.780%
Target Compounds						
16) Acetone	4.156	43	8358	55.833	ug/l	Qvalue # 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
Data File : VD073200.D
Acq On : 20 May 2022 23:13
Operator : VA/SY
Sample : N2885-02
Misc : 6.11G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P004-SS002-0612-01

Quant Time: May 21 00:18:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

