

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073214.D
 Acq On : 21 May 2022 07:18
 Operator : VA/SY
 Sample : N2885-10
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P008-SS002-0006-01

Quant Time: May 22 00:00:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	314135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	549364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	573486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	265987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	173092	49.859	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.720%
35) Dibromofluoromethane	7.909	113	186469	51.439	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.880%
50) Toluene-d8	10.332	98	677967	50.735	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.480%
62) 4-Bromofluorobenzene	12.620	95	270716	57.727	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	115.460%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.973	56	9834	1.621	ug/l #	29

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
Data File : VD073214.D
Acq On : 21 May 2022 07:18
Operator : VA/SY
Sample : N2885-10
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P008-SS002-0006-01

Quant Time: May 22 00:00:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

