

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073216.D
 Acq On : 21 May 2022 08:16
 Operator : VA/SY
 Sample : N2885-12
 Misc : 5.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P008-SS002-1218-01

Quant Time: May 22 00:00:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	311857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	553653	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	583996	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	272171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	177640	51.543	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.080%
35) Dibromofluoromethane	7.908	113	186569	51.068	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.140%
50) Toluene-d8	10.332	98	680748	50.548	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.100%
62) 4-Bromofluorobenzene	12.620	95	262159	55.469	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.940%
Target Compounds						
31) Cyclohexane	7.967	56	8262	1.372	ug/l #	3

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073216.D
 Acq On : 21 May 2022 08:16
 Operator : VA/SY
 Sample : N2885-12
 Misc : 5.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P008-SS002-1218-01

Quant Time: May 22 00:00:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

