

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD052022\  
 Data File : VD073207.D  
 Acq On : 21 May 2022 02:34  
 Operator : VA/SY  
 Sample : VSTDCCC050  
 Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 05/24/2022  
 Supervised By :Mahesh Dadoda 05/24/2022

Quant Time: May 21 03:54:56 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D051922S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 19 15:13:42 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 7.973          | 168  | 301476              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.855          | 114  | 530258              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.638         | 117  | 514078              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.561         | 152  | 251662              | 50.000  | ug/l   | 0.00     |
|                              |                |      |                     |         |        |          |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.326          | 65   | 181375              | 54.439  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 50 - 163 |      | Recovery = 108.880% |         |        |          |
| 35) Dibromofluoromethane     | 7.908          | 113  | 182437              | 52.140  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 54 - 147 |      | Recovery = 104.280% |         |        |          |
| 50) Toluene-d8               | 10.332         | 98   | 673871              | 52.246  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 49 - 140 |      | Recovery = 104.500% |         |        |          |
| 62) 4-Bromofluorobenzene     | 12.620         | 95   | 243729              | 53.845  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 25 - 144 |      | Recovery = 107.680% |         |        |          |
|                              |                |      |                     |         |        |          |
| Target Compounds             |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.991          | 85   | 115798              | 38.440  | ug/l   | 99       |
| 3) Chloromethane             | 2.209          | 50   | 140420              | 41.223  | ug/l   | 99       |
| 4) Vinyl Chloride            | 2.344          | 62   | 127632              | 40.976  | ug/l   | 98       |
| 5) Bromomethane              | 2.762          | 94   | 98711               | 49.383  | ug/l   | 98       |
| 6) Chloroethane              | 2.915          | 64   | 88396               | 46.164  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 3.267          | 101  | 242651              | 44.057  | ug/l   | 96       |
| 8) Diethyl Ether             | 3.709          | 74   | 75297               | 49.752  | ug/l   | 95       |
| 9) 1,1,2-Trichlorotrifluo... | 4.091          | 101  | 144666              | 44.022  | ug/l   | 97       |
| 10) Methyl Iodide            | 4.297          | 142  | 157602              | 44.451  | ug/l   | 99       |
| 11) Tert butyl alcohol       | 5.220          | 59   | 62582               | 121.067 | ug/l # | 95       |
| 12) 1,1-Dichloroethene       | 4.062          | 96   | 137244              | 45.299  | ug/l   | 92       |
| 13) Acrolein                 | 3.926          | 56   | 73596               | 368.971 | ug/l   | 98       |
| 14) Allyl chloride           | 4.709          | 41   | 231293              | 47.294  | ug/l   | 92       |
| 15) Acrylonitrile            | 5.409          | 53   | 187365              | 270.738 | ug/l   | 98       |
| 16) Acetone                  | 4.156          | 43   | 135255              | 234.374 | ug/l   | 91       |
| 17) Carbon Disulfide         | 4.403          | 76   | 424384              | 42.254  | ug/l   | 100      |
| 18) Methyl Acetate           | 4.714          | 43   | 87038               | 54.239  | ug/l   | 94       |
| 19) Methyl tert-butyl Ether  | 5.473          | 73   | 369804              | 53.578  | ug/l   | 98       |
| 20) Methylene Chloride       | 4.961          | 84   | 211803              | 58.545  | ug/l   | 84       |
| 21) trans-1,2-Dichloroethene | 5.467          | 96   | 176511              | 48.184  | ug/l   | 86       |
| 22) Diisopropyl ether        | 6.361          | 45   | 545367              | 53.197  | ug/l   | 92       |
| 23) Vinyl Acetate            | 6.303          | 43   | 1475882             | 269.893 | ug/l # | 94       |
| 24) 1,1-Dichloroethane       | 6.256          | 63   | 326639              | 49.637  | ug/l   | 99       |
| 25) 2-Butanone               | 7.214          | 43   | 223358              | 264.546 | ug/l   | 94       |
| 26) 2,2-Dichloropropane      | 7.203          | 77   | 235544              | 41.972  | ug/l   | 95       |
| 27) cis-1,2-Dichloroethene   | 7.208          | 96   | 202376              | 49.928  | ug/l   | 90       |
| 28) Bromochloromethane       | 7.538          | 49   | 119724              | 52.810  | ug/l   | 84       |
| 29) Tetrahydrofuran          | 7.555          | 42   | 146729              | 268.763 | ug/l   | 91       |
| 30) Chloroform               | 7.703          | 83   | 349478              | 50.870  | ug/l   | 96       |
| 31) Cyclohexane              | 7.979          | 56   | 243736              | 41.863  | ug/l   | 91       |
| 32) 1,1,1-Trichloroethane    | 7.903          | 97   | 282788              | 47.063  | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 8.108          | 75   | 238297              | 45.283  | ug/l   | 96       |
| 37) Ethyl Acetate            | 7.285          | 43   | 102586              | 49.353  | ug/l # | 94       |
| 38) Carbon Tetrachloride     | 8.091          | 117  | 242630              | 45.484  | ug/l   | 99       |
| 39) Methylcyclohexane        | 9.350          | 83   | 250657              | 42.740  | ug/l   | 99       |
| 40) Benzene                  | 8.344          | 78   | 731643              | 48.241  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD052022\  
 Data File : VD073207.D  
 Acq On : 21 May 2022 02:34  
 Operator : VA/SY  
 Sample : VSTDCCC050  
 Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 05/24/2022  
 Supervised By :Mahesh Dadoda 05/24/2022

Quant Time: May 21 03:54:56 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D051922S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 19 15:13:42 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 7.520  | 41   | 54763    | 45.219   | ug/l # | 83       |
| 42) 1,2-Dichloroethane        | 8.414  | 62   | 213724   | 50.645   | ug/l   | 96       |
| 43) Isopropyl Acetate         | 8.450  | 43   | 198298   | 51.612   | ug/l # | 91       |
| 44) Trichloroethene           | 9.108  | 130  | 191279   | 47.295   | ug/l   | 95       |
| 45) 1,2-Dichloropropane       | 9.379  | 63   | 191174   | 50.052   | ug/l   | 97       |
| 46) Dibromomethane            | 9.467  | 93   | 106430   | 50.736   | ug/l   | 98       |
| 47) Bromodichloromethane      | 9.655  | 83   | 265923   | 49.544   | ug/l   | 98       |
| 48) Methyl methacrylate       | 9.449  | 41   | 93525    | 48.384   | ug/l # | 80       |
| 49) 1,4-Dioxane               | 9.455  | 88   | 22330    | 1030.778 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 10.220 | 43   | 546048   | 273.641  | ug/l   | 94       |
| 52) Toluene                   | 10.397 | 92   | 466518   | 48.677   | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.614 | 75   | 238047   | 49.091   | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 10.085 | 75   | 291396   | 50.216   | ug/l # | 89       |
| 55) 1,1,2-Trichloroethane     | 10.791 | 97   | 144439   | 50.918   | ug/l   | 97       |
| 56) Ethyl methacrylate        | 10.649 | 69   | 172316   | 54.635   | ug/l   | 86       |
| 57) 1,3-Dichloropropane       | 10.938 | 76   | 249642   | 52.449   | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 9.932  | 63   | 399859   | 265.392  | ug/l   | 95       |
| 59) 2-Hexanone                | 10.973 | 43   | 366125   | 278.295  | ug/l   | 92       |
| 60) Dibromochloromethane      | 11.132 | 129  | 184716   | 52.222   | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 11.238 | 107  | 142007   | 52.389   | ug/l   | 99       |
| 64) Tetrachloroethene         | 10.867 | 164  | 150766   | 44.776   | ug/l   | 96       |
| 65) Chlorobenzene             | 11.661 | 112  | 500014   | 47.412   | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 11.732 | 131  | 197584   | 50.517   | ug/l   | 99       |
| 67) Ethyl Benzene             | 11.738 | 91   | 898858   | 48.552   | ug/l   | 98       |
| 68) m/p-Xylenes               | 11.843 | 106  | 711516   | 98.921   | ug/l   | 97       |
| 69) o-Xylene                  | 12.167 | 106  | 325827   | 49.311   | ug/l   | 98       |
| 70) Styrene                   | 12.185 | 104  | 586327   | 51.576   | ug/l   | 97       |
| 71) Bromoform                 | 12.349 | 173  | 104667   | 51.996   | ug/l # | 100      |
| 73) Isopropylbenzene          | 12.467 | 105  | 860838   | 47.982   | ug/l   | 100      |
| 74) N-amyl acetate            | 12.279 | 43   | 197979   | 50.584   | ug/l   | 95       |
| 75) 1,1,2,2-Tetrachloroethane | 12.714 | 83   | 168118   | 50.809   | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 12.767 | 75   | 137778m  | 55.798   | ug/l   |          |
| 77) Bromobenzene              | 12.749 | 156  | 205216   | 48.592   | ug/l   | 94       |
| 78) n-propylbenzene           | 12.808 | 91   | 1062819  | 47.188   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 12.896 | 91   | 621023   | 47.754   | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 12.949 | 105  | 739909   | 48.686   | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 12.514 | 75   | 51106    | 49.687   | ug/l   | 87       |
| 82) 4-Chlorotoluene           | 12.990 | 91   | 654266   | 47.998   | ug/l   | 99       |
| 83) tert-Butylbenzene         | 13.208 | 119  | 596192   | 46.547   | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 13.255 | 105  | 741154   | 49.365   | ug/l   | 99       |
| 85) sec-Butylbenzene          | 13.385 | 105  | 914282   | 46.748   | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 13.502 | 119  | 759548   | 47.643   | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.502 | 146  | 407876   | 47.914   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 13.579 | 146  | 405044   | 47.136   | ug/l   | 98       |
| 89) n-Butylbenzene            | 13.826 | 91   | 705383   | 46.436   | ug/l   | 100      |
| 90) Hexachloroethane          | 14.096 | 117  | 156057   | 47.308   | ug/l   | 100      |
| 91) 1,2-Dichlorobenzene       | 13.873 | 146  | 362422   | 49.206   | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.490 | 75   | 24866    | 49.073   | ug/l   | 90       |
| 93) 1,2,4-Trichlorobenzene    | 15.143 | 180  | 202400   | 48.403   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.249 | 225  | 104768   | 44.462   | ug/l   | 99       |
| 95) Naphthalene               | 15.384 | 128  | 391893   | 47.757   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.573 | 180  | 184115   | 49.666   | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD052022\  
Data File : VD073207.D  
Acq On : 21 May 2022 02:34  
Operator : VA/SY  
Sample : VSTDCCC050  
Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022

Quant Time: May 21 03:54:56 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D051922S.M  
Quant Title : SW846 8260  
QLast Update : Thu May 19 15:13:42 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD052022\  
Data File : VD073207.D  
Acq On : 21 May 2022 02:34  
Operator : VA/SY  
Sample : VSTDCCC050  
Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
ALS Vial : 22 Sample Multiplier: 1


Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC

Quant Time: May 21 03:54:56 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D051922S.M  
Quant Title : SW846 8260  
QLast Update : Thu May 19 15:13:42 2022  
Response via : Initial Calibration

Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.973 min Scan# 1029  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:168 Resp: 301470  
Ion Ratio Lower Upper  
168 100  
99 62.4 49.8 74.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#2  
Dichlorodifluoromethane  
Concen: 38.440 ug/l  
RT: 1.991 min Scan# 12  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion: 85 Resp: 115798  
Ion Ratio Lower Upper  
85 100  
87 32.1 15.7 47.0





#3  
Chloromethane  
Concen: 41.223 ug/l  
RT: 2.209 min Scan# 49  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 50 Resp: 140420  
Ion Ratio Lower Upper  
50 100  
52 33.8 27.4 41.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#4  
Vinyl Chloride  
Concen: 40.976 ug/l  
RT: 2.344 min Scan# 72  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 62 Resp: 127632  
Ion Ratio Lower Upper  
62 100  
64 30.8 25.4 38.0





#5  
Bromomethane  
Concen: 49.383 ug/l  
RT: 2.762 min Scan# 144  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 94 Resp: 9871  
Ion Ratio Lower Upper  
94 100  
96 94.6 77.6 116.4




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#6  
Chloroethane  
Concen: 46.164 ug/l  
RT: 2.915 min Scan# 169  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion: 64 Resp: 88396  
Ion Ratio Lower Upper  
64 100  
66 31.8 25.9 38.9





#7  
Trichlorofluoromethane  
Concen: 44.057 ug/l  
RT: 3.267 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:101 Resp: 24265  
Ion Ratio Lower Upper  
101 100  
103 65.6 50.1 75.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#8  
Diethyl Ether  
Concen: 49.752 ug/l  
RT: 3.709 min Scan# 304  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Tgt Ion: 74 Resp: 75297  
Ion Ratio Lower Upper  
74 100  
45 99.8 52.6 157.8









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 44.022 ug/l  
RT: 4.091 min Scan# 30  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion:101 Resp: 144660  
Ion Ratio Lower Upper  
101 100  
85 42.5 36.4 54.6  
151 72.7 57.3 85.9




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#10  
Methyl Iodide  
Concen: 44.451 ug/l  
RT: 4.297 min Scan# 404  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:142 Resp: 157602  
Ion Ratio Lower Upper  
142 100  
127 46.1 37.3 55.9  
141 14.6 11.7 17.5





#11  
Tert butyl alcohol  
Concen: 121.067 ug/l  
RT: 5.220 min Scan# 50  
Delta R.T. 0.011 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 59 Resp: 6258  
Ion Ratio Lower Upper  
59 100  
57 7.7 7.8 11.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#12  
1,1-Dichloroethene  
Concen: 45.299 ug/l  
RT: 4.062 min Scan# 364  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 96 Resp: 137244  
Ion Ratio Lower Upper  
96 100  
61 159.3 138.5 207.7  
98 62.7 48.6 73.0





#13  
Acrolein  
Concen: 368.971 ug/l  
RT: 3.926 min Scan# 341  
Delta R.T. 0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 56 Resp: 73590  
Ion Ratio Lower Upper  
56 100  
55 69.8 57.1 85.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#14  
Allyl chloride  
Concen: 47.294 ug/l  
RT: 4.709 min Scan# 474  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion: 41 Resp: 231293  
Ion Ratio Lower Upper  
41 100  
39 67.8 59.3 88.9  
76 37.6 25.7 38.5





#15  
Acrylonitrile  
Concen: 270.738 ug/l  
RT: 5.409 min Scan# 51  
Delta R.T. -0.005 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 53 Resp: 18736  
Ion Ratio Lower Upper  
53 100  
52 81.6 67.2 100.8  
51 37.7 30.3 45.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#16  
Acetone  
Concen: 234.374 ug/l  
RT: 4.156 min Scan# 380  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion: 43 Resp: 135255  
Ion Ratio Lower Upper  
43 100  
58 32.4 22.1 33.1





#17  
Carbon Disulfide  
Concen: 42.254 ug/l  
RT: 4.403 min Scan# 41  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 76 Resp: 424384  
Ion Ratio Lower Upper  
76 100  
78 9.6 7.5 11.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#18  
Methyl Acetate  
Concen: 54.239 ug/l  
RT: 4.714 min Scan# 475  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 43 Resp: 87038  
Ion Ratio Lower Upper  
43 100  
74 23.9 16.7 25.1





#19  
Methyl tert-butyl Ether  
Concen: 53.578 ug/l  
RT: 5.473 min Scan# 604  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 73 Resp: 369804  
Ion Ratio Lower Upper  
73 100  
57 23.7 18.0 27.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#20  
Methylene Chloride  
Concen: 58.545 ug/l  
RT: 4.961 min Scan# 517  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Tgt Ion: 84 Resp: 211803  
Ion Ratio Lower Upper  
84 100  
49 114.2 112.8 169.2  
51 32.9 32.6 48.8  
86 64.6 47.9 71.9






#21  
trans-1,2-Dichloroethene  
Concen: 48.184 ug/l  
RT: 5.467 min Scan# 603  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :

MSVOA\_D

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 96 Resp: 17651  
Ion Ratio Lower Upper  
96 100  
61 123.5 117.9 176.9  
98 59.7 49.4 74.2

Manual Integrations


APPROVED

Reviewed By :Krupa Patel 05/24/2022



Supervised By :Mahesh Dadoda 05/24/2022




#22  
Diisopropyl ether  
Concen: 53.197 ug/l  
RT: 6.361 min Scan# 755  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 45 Resp: 545367  
Ion Ratio Lower Upper  
45 100  
43 52.2 46.6 70.0  
87 27.4 18.4 27.6  
59 11.0 8.6 13.0





#23  
 Vinyl Acetate  
 Concen: 269.893 ug/l  
 RT: 6.303 min Scan# 744  
 Delta R.T. 0.006 min  
 Lab File: VD073207.D  
 Acq: 21 May 2022 02:34




Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDCCC050EC



Tgt Ion: 43 Resp: 147588  
 Ion Ratio Lower Upper  
 43 100  
 86 10.9 7.1 10.7




Manual Integrations  
 APPROVED


Reviewed By :Krupa Patel 05/24/2022  
 Supervised By :Mahesh Dadoda 05/24/2022



#24  
 1,1-Dichloroethane  
 Concen: 49.637 ug/l  
 RT: 6.256 min Scan# 737  
 Delta R.T. -0.005 min  
 Lab File: VD073207.D  
 Acq: 21 May 2022 02:34


Tgt Ion: 63 Resp: 326639  
 Ion Ratio Lower Upper  
 63 100  
 98 6.7 3.6 10.9  
 100 4.5 2.1 6.5





#25  
2-Butanone  
Concen: 264.546 ug/l  
RT: 7.214 min Scan# 900  
Delta R.T. 0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 223358  
Ion Ratio Lower Upper  
43 100  
72 26.4 18.6 27.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#26  
2,2-Dichloropropane  
Concen: 41.972 ug/l  
RT: 7.203 min Scan# 898  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 77 Resp: 235544  
Ion Ratio Lower Upper  
77 100  
97 24.1 10.8 32.4





#27  
cis-1,2-Dichloroethene  
Concen: 49.928 ug/l  
RT: 7.208 min Scan# 899  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 96 Resp: 202370  
Ion Ratio Lower Upper  
96 100  
61 138.5 0.0 313.8  
98 64.6 0.0 128.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#28  
Bromochloromethane  
Concen: 52.810 ug/l  
RT: 7.538 min Scan# 955  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 49 Resp: 119724  
Ion Ratio Lower Upper  
49 100  
129 2.0 0.0 4.0  
130 78.8 52.6 79.0





#29  
Tetrahydrofuran  
Concen: 268.763 ug/l  
RT: 7.555 min Scan# 91  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 42 Resp: 146729  
Ion Ratio Lower Upper  
42 100  
72 43.3 30.7 46.1  
71 41.7 28.6 43.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#30  
Chloroform  
Concen: 50.870 ug/l  
RT: 7.703 min Scan# 983  
Delta R.T. -0.005 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 83 Resp: 349478  
Ion Ratio Lower Upper  
83 100  
85 62.6 52.7 79.1





#31  
Cyclohexane  
Concen: 41.863 ug/l  
RT: 7.979 min Scan# 1030  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 56 Resp: 243730  
Ion Ratio Lower Upper  
56 100  
69 32.6 25.4 38.2  
84 89.6 63.0 94.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#32  
1,1,1-Trichloroethane  
Concen: 47.063 ug/l  
RT: 7.903 min Scan# 1017  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Tgt Ion: 97 Resp: 282788  
Ion Ratio Lower Upper  
97 100  
99 64.8 52.4 78.6  
61 44.2 38.0 57.0






#33  
1,2-Dichloroethane-d4  
Concen: 54.439 ug/l  
RT: 8.326 min Scan# 1089  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :

MSVOA\_D

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 65 Resp: 181375  
Ion Ratio Lower Upper  
65 100  
67 52.4 0.0 101.4

Manual Integrations




APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.855 min Scan# 1179  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:114 Resp: 530258  
Ion Ratio Lower Upper  
114 100  
63 21.2 0.0 46.6  
88 15.6 0.0 33.2





#35  
Dibromofluoromethane  
Concen: 52.140 ug/l  
RT: 7.908 min Scan# 1019  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:113 Resp: 18243  
Ion Ratio Lower Upper  
113 100  
111 103.0 83.8 125.6  
192 18.2 14.6 21.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#36  
1,1-Dichloropropene  
Concen: 45.283 ug/l  
RT: 8.108 min Scan# 1052  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 75 Resp: 238297  
Ion Ratio Lower Upper  
75 100  
110 35.7 16.4 49.1  
77 31.8 24.2 36.2





#37  
Ethyl Acetate  
Concen: 49.353 ug/l  
RT: 7.285 min Scan# 913  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 102580  
Ion Ratio Lower Upper  
43 100  
61 14.1 9.8 14.8  
70 11.8 6.9 10.3




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#38  
Carbon Tetrachloride  
Concen: 45.484 ug/l  
RT: 8.091 min Scan# 1049  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:117 Resp: 242630  
Ion Ratio Lower Upper  
117 100  
119 95.5 75.8 113.6  
121 31.8 24.5 36.7





#39  
Methylcyclohexane  
Concen: 42.740 ug/l  
RT: 9.350 min Scan# 1263  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 25065  
Ion Ratio Lower Upper  
83 100  
55 79.2 63.2 94.8  
98 49.3 38.2 57.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#40  
Benzene  
Concen: 48.241 ug/l  
RT: 8.344 min Scan# 1092  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 78 Resp: 731643  
Ion Ratio Lower Upper  
78 100  
77 24.5 19.1 28.7





#41  
Methacrylonitrile  
Concen: 45.219 ug/l  
RT: 7.520 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 41 Resp: 5476  
Ion Ratio Lower Upper  
41 100  
39 60.2 44.6 67.0  
67 82.8 48.2 72.4  
52 35.1 22.8 34.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#42  
1,2-Dichloroethane  
Concen: 50.645 ug/l  
RT: 8.414 min Scan# 1104  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 62 Resp: 213724  
Ion Ratio Lower Upper  
62 100  
98 10.4 0.0 17.6





#43  
Isopropyl Acetate  
Concen: 51.612 ug/l  
RT: 8.450 min Scan# 1110  
Delta R.T. 0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 198298  
Ion Ratio Lower Upper  
43 100  
61 33.1 22.4 33.6  
87 14.2 8.7 13.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#44  
Trichloroethene  
Concen: 47.295 ug/l  
RT: 9.108 min Scan# 1222  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:130 Resp: 191279  
Ion Ratio Lower Upper  
130 100  
95 97.4 0.0 205.4





#45  
1,2-Dichloropropane  
Concen: 50.052 ug/l  
RT: 9.379 min Scan# 1268  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 63 Resp: 191174  
Ion Ratio Lower Upper  
63 100  
65 31.2 26.2 39.4




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#46  
Dibromomethane  
Concen: 50.736 ug/l  
RT: 9.467 min Scan# 1283  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion: 93 Resp: 106430  
Ion Ratio Lower Upper  
93 100  
95 82.3 66.7 100.1  
174 92.7 71.6 107.4





#47  
Bromodichloromethane  
Concen: 49.544 ug/l  
RT: 9.655 min Scan# 1315  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 26592  
Ion Ratio Lower Upper  
83 100  
85 65.2 51.4 77.0  
127 9.0 6.1 9.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#48  
Methyl methacrylate  
Concen: 48.384 ug/l  
RT: 9.449 min Scan# 1280  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 41 Resp: 93525  
Ion Ratio Lower Upper  
41 100  
69 94.6 56.7 85.1#  
39 59.7 42.7 64.1





#49  
1,4-Dioxane  
Concen: 1030.778 ug/l  
RT: 9.455 min Scan# 1280  
Delta R.T. 0.005 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 88 Resp: 22330  
Ion Ratio Lower Upper  
88 100  
43 38.5 31.0 46.4  
58 68.6 58.1 87.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#50  
Toluene-d8  
Concen: 52.246 ug/l  
RT: 10.332 min Scan# 1430  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 98 Resp: 673871  
Ion Ratio Lower Upper  
98 100  
100 65.1 52.4 78.6





#51  
4-Methyl-2-Pentanone  
Concen: 273.641 ug/l  
RT: 10.220 min Scan# 1411  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 546048  
Ion Ratio Lower Upper  
43 100  
58 39.5 28.6 43.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#52  
Toluene  
Concen: 48.677 ug/l  
RT: 10.397 min Scan# 1441  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 92 Resp: 466518  
Ion Ratio Lower Upper  
92 100  
91 175.4 139.6 209.4





#53  
t-1,3-Dichloropropene  
Concen: 49.091 ug/l  
RT: 10.614 min Scan# 1478  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 75 Resp: 23804  
Ion Ratio Lower Upper  
75 100  
77 29.5 24.5 36.7

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#54  
cis-1,3-Dichloropropene  
Concen: 50.216 ug/l  
RT: 10.085 min Scan# 1388  
Delta R.T. 0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 75 Resp: 291396  
Ion Ratio Lower Upper  
75 100  
77 31.4 24.8 37.2  
39 45.8 46.8 70.2#





#55  
1,1,2-Trichloroethane  
Concen: 50.918 ug/l  
RT: 10.791 min Scan# 1508  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 97 Resp: 144439  
Ion Ratio Lower Upper  
97 100  
83 87.1 69.0 103.4  
85 58.9 42.6 63.8  
99 65.2 51.7 77.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#56  
Ethyl methacrylate  
Concen: 54.635 ug/l  
RT: 10.649 min Scan# 1484  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Tgt Ion: 69 Resp: 172316  
Ion Ratio Lower Upper  
69 100  
41 64.8 63.0 94.6  
39 36.0 33.8 50.8






#57  
1,3-Dichloropropane  
Concen: 52.449 ug/l  
RT: 10.938 min Scan# 1533  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :

MSVOA\_D

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 76 Resp: 24964  
Ion Ratio Lower Upper  
76 100  
78 32.4 26.0 39.0

Manual Integrations




APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#58  
2-Chloroethyl Vinyl ether  
Concen: 265.392 ug/l  
RT: 9.932 min Scan# 1362  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion: 63 Resp: 399859  
Ion Ratio Lower Upper  
63 100  
106 29.2 21.4 32.0





#59  
2-Hexanone  
Concen: 278.295 ug/l  
RT: 10.973 min Scan# 1539  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 366129  
Ion Ratio Lower Upper  
43 100  
58 55.8 25.3 75.9

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#60  
Dibromochloromethane  
Concen: 52.222 ug/l  
RT: 11.132 min Scan# 1566  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:129 Resp: 184716  
Ion Ratio Lower Upper  
129 100  
127 77.9 39.2 117.6





#61  
1,2-Dibromoethane  
Concen: 52.389 ug/l  
RT: 11.238 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:107 Resp: 14200  
Ion Ratio Lower Upper  
107 100  
109 93.3 74.2 111.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#62  
4-Bromofluorobenzene  
Concen: 53.845 ug/l  
RT: 12.620 min Scan# 1819  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 95 Resp: 243729  
Ion Ratio Lower Upper  
95 100  
174 75.3 0.0 152.8  
176 73.5 0.0 145.4





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.638 min Scan# 1652  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:117 Resp: 514078  
Ion Ratio Lower Upper  
117 100  
82 53.5 45.3 67.9  
119 30.7 26.6 40.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#64  
Tetrachloroethene  
Concen: 44.776 ug/l  
RT: 10.867 min Scan# 1521  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:164 Resp: 150766  
Ion Ratio Lower Upper  
164 100  
166 129.6 98.5 147.7  
129 96.1 74.6 111.8  
131 93.8 72.0 108.0





#65  
Chlorobenzene  
Concen: 47.412 ug/l  
RT: 11.661 min Scan# 1656  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:112 Resp: 500014  
Ion Ratio Lower Upper  
112 100  
114 31.8 25.8 38.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 50.517 ug/l  
RT: 11.732 min Scan# 1668  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:131 Resp: 197584  
Ion Ratio Lower Upper  
131 100  
133 94.5 48.1 144.3  
119 64.7 31.9 95.9





#67  
Ethyl Benzene  
Concen: 48.552 ug/l  
RT: 11.738 min Scan# 1669  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 91 Resp: 898858  
Ion Ratio Lower Upper  
91 100  
106 29.4 24.2 36.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#68  
m/p-Xylenes  
Concen: 98.921 ug/l  
RT: 11.843 min Scan# 1687  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:106 Resp: 711516  
Ion Ratio Lower Upper  
106 100  
91 200.5 163.7 245.5





#69  
o-Xylene  
Concen: 49.311 ug/l  
RT: 12.167 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:106 Resp: 325827  
Ion Ratio Lower Upper  
106 100  
91 217.3 110.4 331.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#70  
Styrene  
Concen: 51.576 ug/l  
RT: 12.185 min Scan# 1745  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:104 Resp: 586327  
Ion Ratio Lower Upper  
104 100  
78 49.2 41.8 62.8  
103 53.9 44.2 66.4





#71  
Bromoform  
Concen: 51.996 ug/l  
RT: 12.349 min Scan# 1773  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:173 Resp: 10466  
Ion Ratio Lower Upper  
173 100  
175 47.7 23.9 71.8  
254 0.0 0.0 0.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.561 min Scan# 1979  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:152 Resp: 251662  
Ion Ratio Lower Upper  
152 100  
115 60.2 31.3 93.8  
150 173.3 0.0 348.6





#73  
Isopropylbenzene  
Concen: 47.982 ug/l  
RT: 12.467 min Scan# 1793  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:105 Resp: 860838  
Ion Ratio Lower Upper  
105 100  
120 25.7 13.0 38.9

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#74  
N-ethyl acetate  
Concen: 50.584 ug/l  
RT: 12.279 min Scan# 1761  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 43 Resp: 197979  
Ion Ratio Lower Upper  
43 100  
70 43.9 31.3 46.9  
55 26.6 20.0 30.0  
61 23.6 17.8 26.8





#75  
1,1,2,2-Tetrachloroethane  
Concen: 50.809 ug/l  
RT: 12.714 min Scan# 1835  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 168118  
Ion Ratio Lower Upper  
83 100  
131 10.0 5.0 14.9  
85 63.7 32.6 98.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#76  
1,2,3-Trichloropropane  
Concen: 55.798 ug/l m  
RT: 12.767 min Scan# 1844  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 75 Resp: 137778  
Ion Ratio Lower Upper  
75 100  
77 284.5 0.0 0.0#





#77  
Bromobenzene  
Concen: 48.592 ug/l  
RT: 12.749 min Scan# 1841  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:156 Resp: 205210  
Ion Ratio Lower Upper  
156 100  
77 191.0 102.8 308.3  
158 97.4 48.7 146.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#78  
n-propylbenzene  
Concen: 47.188 ug/l  
RT: 12.808 min Scan# 1851  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Tgt Ion: 91 Resp: 1062819  
Ion Ratio Lower Upper  
91 100  
120 22.5 10.9 32.9









#79  
2-Chlorotoluene  
Concen: 47.754 ug/l  
RT: 12.896 min Scan# 1866  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 91 Resp: 621023  
Ion Ratio Lower Upper  
91 100  
126 33.7 16.2 48.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#80  
1,3,5-Trimethylbenzene  
Concen: 48.686 ug/l  
RT: 12.949 min Scan# 1875  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:105 Resp: 739909  
Ion Ratio Lower Upper  
105 100  
120 48.9 24.3 72.8





#81  
trans-1,4-Dichloro-2-butene  
Concen: 49.687 ug/l  
RT: 12.514 min Scan# 1801  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 75 Resp: 51100  
Ion Ratio Lower Upper  
75 100  
53 89.3 86.5 129.7  
89 41.7 33.9 50.9

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#82  
4-Chlorotoluene  
Concen: 47.998 ug/l  
RT: 12.990 min Scan# 1882  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Tgt Ion: 91 Resp: 654266  
Ion Ratio Lower Upper  
91 100  
126 33.6 16.5 49.5





#83  
tert-Butylbenzene  
Concen: 46.547 ug/l  
RT: 13.208 min Scan# 1919  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :

MSVOA\_D

ClientSampleId :





VSTDCCC050EC

Tgt Ion:119 Resp: 596192  
Ion Ratio Lower Upper  
119 100  
91 68.7 34.8 104.5  
134 27.9 13.3 39.9

Manual Integrations




APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#84  
1,2,4-Trimethylbenzene  
Concen: 49.365 ug/l  
RT: 13.255 min Scan# 1927  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:105 Resp: 741154  
Ion Ratio Lower Upper  
105 100  
120 45.1 22.3 66.8





#85  
 sec-Butylbenzene  
 Concen: 46.748 ug/l  
 RT: 13.385 min Scan# 1949  
 Delta R.T. -0.006 min  
 Lab File: VD073207.D  
 Acq: 21 May 2022 02:34




Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDCCC050EC




Tgt Ion:105 Resp: 91428  
 Ion Ratio Lower Upper  
 105 100  
 134 19.4 9.8 29.3

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 05/24/2022  
 Supervised By :Mahesh Dadoda 05/24/2022




#86  
 p-Isopropyltoluene  
 Concen: 47.643 ug/l  
 RT: 13.502 min Scan# 1969  
 Delta R.T. 0.000 min  
 Lab File: VD073207.D  
 Acq: 21 May 2022 02:34




Tgt Ion:119 Resp: 759548  
 Ion Ratio Lower Upper  
 119 100  
 134 26.4 12.9 38.6  
 91 24.1 12.1 36.3





#87  
1,3-Dichlorobenzene  
Concen: 47.914 ug/l  
RT: 13.502 min Scan# 1969  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:146 Resp: 407870  
Ion Ratio Lower Upper  
146 100  
111 40.5 20.4 61.3  
148 64.7 31.9 95.7

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#88  
1,4-Dichlorobenzene  
Concen: 47.136 ug/l  
RT: 13.579 min Scan# 1982  
Delta R.T. -0.006 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Tgt Ion:146 Resp: 405044  
Ion Ratio Lower Upper  
146 100  
111 39.1 20.3 60.9  
148 65.1 31.7 95.1






#89  
n-Butylbenzene  
Concen: 46.436 ug/l  
RT: 13.826 min Scan# 2024  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34

Instrument :

MSVOA\_D

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 91 Resp: 70538  
Ion Ratio Lower Upper  
91 100  
92 53.3 26.6 79.8  
134 25.2 12.3 36.9

Manual Integrations




APPROVED


Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#90  
Hexachloroethane  
Concen: 47.308 ug/l  
RT: 14.096 min Scan# 2070  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34


Tgt Ion:117 Resp: 156057  
Ion Ratio Lower Upper  
117 100  
201 66.5 33.4 100.2





#91  
1,2-Dichlorobenzene  
Concen: 49.206 ug/l  
RT: 13.873 min Scan# 2032  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:146 Resp: 36242  
Ion Ratio Lower Upper  
146 100  
111 42.6 21.1 63.3  
148 64.2 31.7 95.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 49.073 ug/l  
RT: 14.490 min Scan# 2137  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion: 75 Resp: 24866  
Ion Ratio Lower Upper  
75 100  
155 85.3 36.8 110.3  
157 102.7 48.4 145.0





#93  
1,2,4-Trichlorobenzene  
Concen: 48.403 ug/l  
RT: 15.143 min Scan# 2248  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:180 Resp: 202400  
Ion Ratio Lower Upper  
180 100  
182 93.4 47.3 142.0  
145 32.3 16.2 48.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022




#94  
Hexachlorobutadiene  
Concen: 44.462 ug/l  
RT: 15.249 min Scan# 2266  
Delta R.T. 0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Tgt Ion:225 Resp: 104768  
Ion Ratio Lower Upper  
225 100  
223 62.7 31.1 93.2  
227 63.0 32.4 97.0





#95  
Naphthalene  
Concen: 47.757 ug/l  
RT: 15.384 min Scan# 2289  
Delta R.T. -0.001 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:128 Resp: 39189  
Ion Ratio Lower Upper  
128 100  
127 13.2 10.3 15.5  
129 11.1 8.4 12.6

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 05/24/2022  
Supervised By :Mahesh Dadoda 05/24/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 49.666 ug/l  
RT: 15.573 min Scan# 2321  
Delta R.T. -0.000 min  
Lab File: VD073207.D  
Acq: 21 May 2022 02:34



Tgt Ion:180 Resp: 184115  
Ion Ratio Lower Upper  
180 100  
182 95.3 47.6 142.9  
145 33.2 17.1 51.3

