



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073187.D
 Acq On : 20 May 2022 16:30
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Title : SW846 8260
 Last Update : Thu May 19 15:13:42 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	29693	PASS
75	95	30	60	53.0	79293	PASS
95	95	100	100	100.0	149589	PASS
96	95	5	9	6.8	10201	PASS
173	174	0.00	2	1.3	1573	PASS
174	95	50	100	79.3	118579	PASS
175	174	5	9	7.4	8814	PASS
176	174	95	101	95.1	112747	PASS
177	176	5	9	6.8	7720	PASS