

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052121\
 Data File : VD069236.D
 Acq On : 21 May 2021 14:00
 Operator : VA/SY
 Sample : M2472-10
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW8-GW-888-890

Manual Integrations
 APPROVED

MMDadoda
 5/24/2021 3:43:56 PM

Quant Time: May 22 00:33:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	43833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	79194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	83985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	31180	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	47134	107.26	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 214.52%#			
35) Dibromofluoromethane	7.920	113	38758	75.20	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 150.40%#			
50) Toluene-d8	10.338	98	92953	49.09	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 98.18%			
62) 4-Bromofluorobenzene	12.626	95	27978	45.92	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 91.84%			
Target Compounds					Qvalue	
16) Acetone	4.150	43	4945m	73.48	ug/l	
20) Methylene Chloride	4.956	84	8377	7.60	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051212\
Data File : VD069236.D
Acq On : 21 May 2021 14:00
Operator : VA/SY
Sample : M2472-10
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW8-GW-888-890

Manual Integrations
APPROVED

MMDadoda
5/24/2021 3:43:56 PM

Quant Time: May 22 00:33:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

