

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052121\
 Data File : VD069242.D
 Acq On : 21 May 2021 16:55
 Operator : VA/SY
 Sample : M2486-01
 Misc : 6.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-052121

Quant Time: May 22 00:35:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	122344	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	214262	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	200007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	86683	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	103538	84.42	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	168.84%#
35) Dibromofluoromethane	7.914	113	94474	67.75	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	135.50%
50) Toluene-d8	10.338	98	267178	52.15	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.30%
62) 4-Bromofluorobenzene	12.626	95	78578	47.66	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.32%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052121\
Data File : VD069242.D
Acq On : 21 May 2021 16:55
Operator : VA/SY
Sample : M2486-01
Misc : 6.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-052121

Quant Time: May 22 00:35:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

