


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD052219\
 Data File : VD062496.D
 Acq On : 22 May 2019 9:49
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D052019S.M
 Title : SW846 8260
 Last Update : Tue May 21 03:32:39 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	19565	PASS
75	95	30	60	45.3	41144	PASS
95	95	100	100	100.0	90736	PASS
96	95	5	9	6.4	5777	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.4	80197	PASS
175	174	5	9	7.4	5905	PASS
176	174	95	101	100.9	80906	PASS
177	176	5	9	6.5	5292	PASS