

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052223\
 Data File : VD076181.D
 Acq On : 22 May 2023 17:52
 Operator : KP/SY
 Sample : 02840-02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1

Quant Time: May 23 01:25:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3112	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	5562	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	2005	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	1950	64.834	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.660%
35) Dibromofluoromethane	7.798	113	2090	54.363	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.720%
50) Toluene-d8	10.269	98	4593	30.646	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	61.300%
62) 4-Bromofluorobenzene	12.575	95	1707	41.002	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052223\
Data File : VD076181.D
Acq On : 22 May 2023 17:52
Operator : KP/SY
Sample : 02840-02
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1

Quant Time: May 23 01:25:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

