

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052223\
 Data File : VD076182.D
 Acq On : 22 May 2023 18:22
 Operator : KP/SY
 Sample : 02840-04
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2

Quant Time: May 23 01:25:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.858	168	216	50.000	ug/l	#-0.02
34) 1,4-Difluorobenzene	8.781	114	858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	61	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.52

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	7.816	113	121	20.402	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	40.800%#
50) Toluene-d8	10.269	98	404	17.475	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	34.940%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	30 - 143	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052223\
Data File : VD076182.D
Acq On : 22 May 2023 18:22
Operator : KP/SY
Sample : 02840-04
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2

Quant Time: May 23 01:25:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

