

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069255.D
 Acq On : 24 May 2021 14:53
 Operator : VA/SY
 Sample : M2498-03
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW8-GW-958-960

Quant Time: May 25 00:55:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	145063	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	273055	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	249264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	107600	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	98998	68.07	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 136.14%			
35) Dibromofluoromethane	7.914	113	104281	58.69	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 117.38%			
50) Toluene-d8	10.338	98	336860	51.59	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 103.18%			
62) 4-Bromofluorobenzene	12.626	95	103710	49.36	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 98.72%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	13285	59.65	ug/l	97
17) Carbon Disulfide	4.414	76	5927	1.23	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
Data File : VD069255.D
Acq On : 24 May 2021 14:53
Operator : VA/SY
Sample : M2498-03
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW8-GW-958-960

Quant Time: May 25 00:55:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

