

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069256.D
 Acq On : 24 May 2021 15:21
 Operator : VA/SY
 Sample : M2479-01
 Misc : 5.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SUBBASE-6-1-D

Quant Time: May 25 00:56:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	226371	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	420406	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	379260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	144142	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	138082	60.85	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 121.70%		
35) Dibromofluoromethane	7.914	113	147926	54.07	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 108.14%		
50) Toluene-d8	10.338	98	506317	50.37	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 100.74%		
62) 4-Bromofluorobenzene	12.626	95	135707	41.95	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 83.90%		
Target Compounds						Qvalue
20) Methylene Chloride	4.967	84	44451	8.11	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
Data File : VD069256.D
Acq On : 24 May 2021 15:21
Operator : VA/SY
Sample : M2479-01
Misc : 5.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SUBBASE-6-1-D

Quant Time: May 25 00:56:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

