

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069258.D
 Acq On : 24 May 2021 16:18
 Operator : VA/SY
 Sample : M2479-03
 Misc : 5.15G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GROUND-MATERIAL-1-17-D

Quant Time: May 25 00:56:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	195047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	372088	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	350984	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	148539	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	122108	62.45	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.90%
35) Dibromofluoromethane	7.914	113	134512	55.55	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.10%
50) Toluene-d8	10.338	98	471727	53.02	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.04%
62) 4-Bromofluorobenzene	12.632	95	138413	48.35	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.70%
Target Compounds						
20) Methylene Chloride	4.967	84	48492	13.22	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
Data File : VD069258.D
Acq On : 24 May 2021 16:18
Operator : VA/SY
Sample : M2479-03
Misc : 5.15G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GROUND-MATERIAL-1-17-D

Quant Time: May 25 00:56:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

