

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069262.D
 Acq On : 24 May 2021 18:13
 Operator : VA/SY
 Sample : M2489-01
 Misc : 6.58G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 359

Quant Time: May 25 00:57:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	170393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	329068	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	318119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	137415	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	123537	72.32	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	144.64%
35) Dibromofluoromethane	7.920	113	124736	58.25	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.50%
50) Toluene-d8	10.338	98	420108	53.39	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.78%
62) 4-Bromofluorobenzene	12.626	95	137358	54.25	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
Data File : VD069262.D
Acq On : 24 May 2021 18:13
Operator : VA/SY
Sample : M2489-01
Misc : 6.58G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
359

Quant Time: May 25 00:57:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

