

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069264.D
 Acq On : 24 May 2021 19:11
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 25 00:58:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.532	0.453	14.8	69	0.00
3 P	Chloromethane	0.611	0.637	-4.3	88	0.00
4 C	Vinyl Chloride	0.805	0.782	2.9#	76	0.00
5 T	Bromomethane	0.606	0.567	6.4	81	0.00
6 T	Chloroethane	0.518	0.577	-11.4	87	0.00
7 T	Trichlorofluoromethane	0.982	0.977	0.5	82	0.00
8 T	Diethyl Ether	0.224	0.264	-17.9	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.545	2.2	78	0.00
10 T	Methyl Iodide	0.531	0.464	12.6	60	0.00
11 T	Tert butyl alcohol	0.034	0.031	8.8	91	0.00
12 CM	1,1-Dichloroethene	0.500	0.472	5.6#	72	0.00
13 T	Acrolein	0.041	0.037	9.8	68	0.00
14 T	Allyl chloride	0.549	0.569	-3.6	80	0.00
15 T	Acrylonitrile	0.084	0.101	-20.2	91	0.00
16 T	Acetone	0.077	0.078	-1.3	76	0.00
17 T	Carbon Disulfide	1.658	1.541	7.1	74	-0.01
18 T	Methyl Acetate	0.187	0.237	-26.7#	96	0.00
19 T	Methyl tert-butyl Ether	0.908	1.066	-17.4	86	0.00
20 T	Methylene Chloride	0.813	0.720	11.4	85	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.608	-2.5	80	0.00
22 T	Diisopropyl ether	1.096	1.285	-17.2	86	0.00
23 T	Vinyl Acetate	0.647	0.786	-21.5	85	0.00
24 P	1,1-Dichloroethane	0.947	1.011	-6.8	84	0.00
25 T	2-Butanone	0.104	0.123	-18.3	88	0.00
26 T	2,2-Dichloropropane	0.885	0.836	5.5	74	0.00
27 T	cis-1,2-Dichloroethene	0.644	0.676	-5.0	83	0.00
28 T	Bromochloromethane	0.338	0.375	-10.9	88	-0.01
29 T	Tetrahydrofuran	0.057	0.073	-28.1#	93	0.00
30 C	Chloroform	1.066	1.149	-7.8#	85	0.00
31 T	Cyclohexane	0.743	0.669	10.0	74	0.00
32 T	1,1,1-Trichloroethane	0.967	0.997	-3.1	81	0.00
33 S	1,2-Dichloroethane-d4	0.501	0.534	-6.6	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.325	0.341	-4.9	90	0.00
36 T	1,1-Dichloropropene	0.486	0.492	-1.2	79	0.00
37 T	Ethyl Acetate	0.156	0.182	-16.7	91	0.00
38 T	Carbon Tetrachloride	0.543	0.519	4.4	77	0.00
39 T	Methylcyclohexane	0.513	0.481	6.2	71	0.00
40 TM	Benzene	1.392	1.452	-4.3	83	0.00
41 T	Methacrylonitrile	0.083	0.090	-8.4	88	0.00
42 TM	1,2-Dichloroethane	0.380	0.418	-10.0	86	0.00
43 T	Isopropyl Acetate	0.284	0.339	-19.4	91	0.00
44 TM	Trichloroethene	0.389	0.391	-0.5	79	0.00
45 C	1,2-Dichloropropane	0.329	0.366	-11.2#	89	0.00
46 T	Dibromomethane	0.196	0.220	-12.2	88	0.00
47 T	Bromodichloromethane	0.500	0.549	-9.8	85	0.00
48 T	Methyl methacrylate	0.131	0.149	-13.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069264.D
 Acq On : 24 May 2021 19:11
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 25 00:58:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	94	0.00
50 S	Toluene-d8	1.196	1.202	-0.5	86	0.00
51 T	4-Methyl-2-Pentanone	0.146	0.186	-27.4#	94	0.00
52 CM	Toluene	0.899	0.954	-6.1#	81	0.00
53 T	t-1,3-Dichloropropene	0.439	0.482	-9.8	84	0.00
54 T	cis-1,3-Dichloropropene	0.525	0.551	-5.0	83	0.00
55 T	1,1,2-Trichloroethane	0.265	0.297	-12.1	88	0.00
56 T	Ethyl methacrylate	0.264	0.322	-22.0	87	0.00
57 T	1,3-Dichloropropane	0.428	0.494	-15.4	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.163	-24.4	92	0.00
59 T	2-Hexanone	0.097	0.124	-27.8#	92	0.00
60 T	Dibromochloromethane	0.340	0.389	-14.4	90	0.00
61 T	1,2-Dibromoethane	0.262	0.292	-11.5	86	0.00
62 S	4-Bromofluorobenzene	0.385	0.415	-7.8	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.347	0.323	6.9	75	0.00
65 PM	Chlorobenzene	1.036	1.062	-2.5	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.406	-3.8	85	0.00
67 C	Ethyl Benzene	1.783	1.865	-4.6#	81	0.00
68 T	m/p-Xylenes	0.699	0.729	-4.3	81	0.00
69 T	o-Xylene	0.622	0.674	-8.4	82	0.00
70 T	Styrene	1.085	1.202	-10.8	85	0.00
71 P	Bromoform	0.203	0.231	-13.8	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.495	3.405	2.6	79	0.00
74 T	N-amyl acetate	0.572	0.649	-13.5	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.633	0.691	-9.2	90	0.00
76 T	1,2,3-Trichloropropane	0.376	0.422	-12.2	87	0.00
77 T	Bromobenzene	0.829	0.855	-3.1	86	0.00
78 T	n-propylbenzene	4.217	4.156	1.4	79	0.00
79 T	2-Chlorotoluene	2.346	2.329	0.7	82	0.00
80 T	1,3,5-Trimethylbenzene	2.881	2.946	-2.3	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.186	0.201	-8.1	90	0.00
82 T	4-Chlorotoluene	2.499	2.476	0.9	82	0.00
83 T	tert-Butylbenzene	2.401	2.374	1.1	79	0.00
84 T	1,2,4-Trimethylbenzene	2.899	2.972	-2.5	82	0.00
85 T	sec-Butylbenzene	3.611	3.478	3.7	77	0.00
86 T	p-Isopropyltoluene	3.172	3.119	1.7	78	0.00
87 T	1,3-Dichlorobenzene	1.678	1.703	-1.5	85	0.00
88 T	1,4-Dichlorobenzene	1.662	1.664	-0.1	86	0.00
89 T	n-Butylbenzene	2.956	2.836	4.1	75	0.00
90 T	Hexachloroethane	0.594	0.570	4.0	81	0.00
91 T	1,2-Dichlorobenzene	1.441	1.462	-1.5	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.098	0.107	-9.2	96	0.00
93 T	1,2,4-Trichlorobenzene	0.883	0.904	-2.4	85	0.00
94 T	Hexachlorobutadiene	0.530	0.506	4.5	81	0.00
95 T	Naphthalene	1.480	1.732	-17.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
Data File : VD069264.D
Acq On : 24 May 2021 19:11
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 25 00:58:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.756	0.821	-8.6	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6