


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052421\
 Data File : VD069249.D
 Acq On : 24 May 2021 11:07
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Title : SW846 8260
 Last Update : Thu May 13 04:01:29 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	12197	PASS
75	95	30	60	51.4	34597	PASS
95	95	100	100	100.0	67315	PASS
96	95	5	9	6.3	4273	PASS
173	174	0.00	2	1.3	690	PASS
174	95	50	100	78.6	52899	PASS
175	174	5	9	8.8	4646	PASS
176	174	95	101	99.3	52544	PASS
177	176	5	9	7.5	3921	PASS