

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052424\
 Data File : VD079032.D
 Acq On : 24 May 2024 22:07
 Operator : RP/MD
 Sample : P2570-03
 Misc : 4.58G/10ml/MSVOA_D/SOIL/VIAL B
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BH6B1

Quant Time: May 24 23:28:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM051724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 24 02:12:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	200770	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	200079	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	79814	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	118236	24.023	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.080%
7) Chloroethane-d5	2.793	69	117604	22.322	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.280%
11) 1,1-Dichloroethene-d2	3.911	65	18684	19.461	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.840%
21) 2-Butanone-d5	6.987	46	17874	25.657	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	51.320%
24) Chloroform-d	7.563	84	95583	15.992	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	63.960%
26) 1,2-Dichloroethane-d4	8.228	65	54472	19.191	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.760%
32) Benzene-d6	8.199	84	198614	19.455	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.800%
36) 1,2-Dichloropropane-d6	9.210	67	60056	19.307	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.240%
41) Toluene-d8	10.269	98	167821	17.834	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	71.320%
43) trans-1,3-Dichloroprop...	10.522	79	24532	17.121	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.480%
47) 2-Hexanone-d5	10.875	63	15232	24.433	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	48.860%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	57970	16.383	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	65.520%
66) 1,2-Dichlorobenzene-d4	13.810	152	53919	19.904	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.600%
Target Compounds						
					Qvalue	
16) Methylene chloride	4.793	84	15467	3.252	ug/L	99
25) Chloroform	7.593	83	7030	0.985	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052424\
Data File : VD079032.D
Acq On : 24 May 2024 22:07
Operator : RP/MD
Sample : P2570-03
Misc : 4.58G/10ml/MSVOA_D/SOIL/VIAL B
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BH6B1

Quant Time: May 24 23:28:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM051724SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 24 02:12:30 2024
Response via : Initial Calibration

