

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052521\
 Data File : VD069267.D
 Acq On : 25 May 2021 11:56
 Operator : VA/SY
 Sample : VD0525SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0525SBL01

Quant Time: May 26 01:03:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	208195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	381160	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	365468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	150765	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	128710	61.67	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.34%
35) Dibromofluoromethane	7.914	113	135502	54.63	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.26%
50) Toluene-d8	10.337	98	492256	54.01	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.02%
62) 4-Bromofluorobenzene	12.626	95	127756	43.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052521\
Data File : VD069267.D
Acq On : 25 May 2021 11:56
Operator : VA/SY
Sample : VD0525SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0525SBL01

Quant Time: May 26 01:03:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

