

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD052620\
 Data File : VD065755.D
 Acq On : 26 May 2020 11:50
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 27 07:36:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D051320S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 13 14:28:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.328	0.293	10.7	76	0.00
3 P	Chloromethane	0.470	0.418	11.1	78	0.00
4 C	Vinyl Chloride	0.509	0.493	3.1#	84	0.00
5 T	Bromomethane	0.336	0.311	7.4	83	0.00
6 T	Chloroethane	0.330	0.329	0.3	85	0.00
7 T	Trichlorofluoromethane	0.704	0.649	7.8	79	0.00
8 T	Diethyl Ether	0.202	0.175	13.4	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.433	0.384	11.3	77	0.00
10 T	Methyl Iodide	0.452	0.395	12.6	72	0.00
11 T	Tert butyl alcohol	0.047	0.025	46.8#	68	0.00
12 CM	1,1-Dichloroethene	0.410	0.358	12.7#	74	0.00
13 T	Acrolein	0.026	0.021	19.2	86	0.00
14 T	Allyl chloride	0.709	0.593	16.4	72	0.00
15 T	Acrylonitrile	0.089	0.078	12.4	76	0.00
16 T	Acetone	0.083	0.100	-20.5#	104	0.00
17 T	Carbon Disulfide	1.296	1.099	15.2	74	0.00
18 T	Methyl Acetate	0.214	0.170	20.6#	77	-0.01
19 T	Methyl tert-butyl Ether	0.957	0.819	14.4	74	0.00
20 T	Methylene Chloride	0.466	0.374	19.7	74	0.00
21 T	trans-1,2-Dichloroethene	0.472	0.409	13.3	75	0.00
22 T	Diisopropyl ether	1.351	1.164	13.8	74	0.00
23 T	Vinyl Acetate	0.794	0.697	12.2	75	0.00
24 P	1,1-Dichloroethane	0.791	0.698	11.8	77	0.00
25 T	2-Butanone	0.118	0.111	5.9	82	0.00
26 T	2,2-Dichloropropane	0.734	0.657	10.5	78	0.00
27 T	cis-1,2-Dichloroethene	0.518	0.443	14.5	75	0.00
28 T	Bromochloromethane	0.325	0.336	-3.4	84	0.00
29 T	Tetrahydrofuran	0.075	0.064	14.7	75	0.00
30 C	Chloroform	0.799	0.725	9.3#	79	0.00
31 T	Cyclohexane	0.800	0.677	15.4	78	0.00
32 T	1,1,1-Trichloroethane	0.749	0.676	9.7	78	0.00
33 S	1,2-Dichloroethane-d4	0.403	0.390	3.2	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.283	0.266	6.0	80	0.00
36 T	1,1-Dichloropropene	0.434	0.384	11.5	77	0.00
37 T	Ethyl Acetate	0.186	0.156	16.1	76	0.00
38 T	Carbon Tetrachloride	0.461	0.416	9.8	79	0.00
39 T	Methylcyclohexane	0.556	0.483	13.1	79	0.00
40 TM	Benzene	1.194	1.058	11.4	77	0.00
41 T	Methacrylonitrile	0.102	0.107	-4.9	93	0.00
42 TM	1,2-Dichloroethane	0.332	0.304	8.4	78	0.00
43 T	Isopropyl Acetate	0.369	0.311	15.7	75	0.00
44 TM	Trichloroethene	0.357	0.306	14.3	74	0.00
45 C	1,2-Dichloropropane	0.294	0.266	9.5#	78	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD052620\
 Data File : VD065755.D
 Acq On : 26 May 2020 11:50
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 27 07:36:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D051320S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 13 14:28:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.159	0.141	11.3	77	0.00
47 T	Bromodichloromethane	0.416	0.373	10.3	78	0.00
48 T	Methyl methacrylate	0.173	0.151	12.7	73	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	75	0.00
50 S	Toluene-d8	1.078	1.005	6.8	80	0.00
51 T	4-Methyl-2-Pentanone	0.176	0.154	12.5	76	0.00
52 CM	Toluene	0.782	0.688	12.0#	76	0.00
53 T	t-1,3-Dichloropropene	0.406	0.353	13.1	75	0.00
54 T	cis-1,3-Dichloropropene	0.483	0.426	11.8	75	0.00
55 T	1,1,2-Trichloroethane	0.216	0.193	10.6	78	0.00
56 T	Ethyl methacrylate	0.282	0.237	16.0	72	0.00
57 T	1,3-Dichloropropane	0.376	0.331	12.0	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.134	-2.3	83	0.00
59 T	2-Hexanone	0.120	0.115	4.2	83	0.00
60 T	Dibromochloromethane	0.290	0.259	10.7	76	0.00
61 T	1,2-Dibromoethane	0.216	0.186	13.9	75	0.00
62 S	4-Bromofluorobenzene	0.365	0.334	8.5	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.346	0.288	16.8	76	0.00
65 PM	Chlorobenzene	0.942	0.800	15.1	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.353	0.295	16.4	75	0.00
67 C	Ethyl Benzene	1.680	1.462	13.0#	78	0.00
68 T	m/p-Xylenes	0.642	0.562	12.5	78	0.00
69 T	o-Xylene	0.597	0.520	12.9	77	0.00
70 T	Styrene	1.019	0.886	13.1	77	0.00
71 P	Bromoform	0.195	0.165	15.4	73	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.357	2.935	12.6	80	0.00
74 T	N-amyl acetate	0.779	0.646	17.1	73	0.00
75 P	1,1,2,2-Tetrachloroethane	0.531	0.461	13.2	78	0.00
76 T	1,2,3-Trichloropropane	0.373	0.354	5.1	79	0.00
77 T	Bromobenzene	0.809	0.676	16.4	75	0.00
78 T	n-propylbenzene	3.921	3.469	11.5	80	0.00
79 T	2-Chlorotoluene	2.170	1.876	13.5	78	0.00
80 T	1,3,5-Trimethylbenzene	2.766	2.422	12.4	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.161	15.3	75	0.00
82 T	4-Chlorotoluene	2.287	2.003	12.4	80	0.00
83 T	tert-Butylbenzene	2.453	2.121	13.5	79	0.00
84 T	1,2,4-Trimethylbenzene	2.728	2.404	11.9	80	0.00
85 T	sec-Butylbenzene	3.337	2.971	11.0	82	0.00
86 T	p-Isopropyltoluene	3.067	2.736	10.8	81	0.00
87 T	1,3-Dichlorobenzene	1.518	1.311	13.6	79	0.00
88 T	1,4-Dichlorobenzene	1.512	1.284	15.1	77	0.00
89 T	n-Butylbenzene	2.868	2.569	10.4	82	0.00

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD052620\
Data File : VD065755.D
Acq On : 26 May 2020 11:50
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 27 07:36:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D051320S.M
Quant Title : SW846 8260
QLast Update : Wed May 13 14:28:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.600	0.528	12.0	80	0.00
91 T	1,2-Dichlorobenzene	1.313	1.114	15.2	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.089	0.073	18.0	78	0.00
93 T	1,2,4-Trichlorobenzene	0.999	0.830	16.9	76	0.00
94 T	Hexachlorobutadiene	0.599	0.504	15.9	78	0.00
95 T	Naphthalene	1.661	1.390	16.3	76	0.00
96 T	1,2,3-Trichlorobenzene	0.851	0.702	17.5	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6