

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052924\
 Data File : VD079076.D
 Acq On : 29 May 2024 20:41
 Operator : RP/MD
 Sample : P2640-13
 Misc : 5.48G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A4C33

Quant Time: May 30 01:15:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu May 30 01:09:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	264464	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	246560	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	91855	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	152581	18.748	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	75.000%	
7) Chloroethane-d5	2.799	69	151441	19.245	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	76.960%	
11) 1,1-Dichloroethene-d2	3.917	65	24433	16.912	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	67.640%	
21) 2-Butanone-d5	6.993	46	14942	24.004	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	48.000%	
24) Chloroform-d	7.569	84	123978	16.766	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	67.080%	
26) 1,2-Dichloroethane-d4	8.222	65	55091	16.150	ug/L	-0.01
Spiked Amount	25.000	Range 70 - 130	Recovery	=	64.600%#	
32) Benzene-d6	8.199	84	252703	18.488	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	73.960%	
36) 1,2-Dichloropropane-d6	9.210	67	71628	18.192	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	72.760%	
41) Toluene-d8	10.269	98	222329	17.372	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	69.480%	
43) trans-1,3-Dichloroprop...	10.528	79	24452	14.775	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	59.120%	
47) 2-Hexanone-d5	10.875	63	10472	20.574	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	41.140%	
56) 1,1,2,2-Tetrachloroeth...	12.645	84	48103	13.793	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	55.160%	
66) 1,2-Dichlorobenzene-d4	13.810	152	59550	18.785	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	75.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052924\
Data File : VD079076.D
Acq On : 29 May 2024 20:41
Operator : RP/MD
Sample : P2640-13
Misc : 5.48G/10ml/MSVOA_D/SOIL/A
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A4C33

Quant Time: May 30 01:15:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu May 30 01:09:48 2024
Response via : Initial Calibration

