

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD053119\
 Data File : VD062564.D
 Acq On : 31 May 2019 15:26
 Operator : VA/AP
 Sample : K3125-21
 Misc : 5.04g/5ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 13903

Quant Time: Jun 01 00:11:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D053019S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 31 02:08:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	423597	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	636644	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.38	117	344218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	79610	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	278515	56.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.16%	
35) Dibromofluoromethane	6.94	113	275078	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
50) Toluene-d8	10.42	98	348677	27.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.86%	
62) 4-Bromofluorobenzene	13.56	95	104719	18.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	36.54%	
Target Compounds						
11) Tert butyl alcohol	4.11	59	12658	59.751	ug/l #	80
17) Carbon Disulfide	3.39	76	12819	1.190	ug/l	96
52) Toluene	10.53	92	420951	49.620	ug/l	99
68) m/p-Xylenes	12.67	106	8687	2.044	ug/l	88
80) 1,3,5-Trimethylbenzene	13.93	105	9483	2.115	ug/l	82
84) 1,2,4-Trimethylbenzene	14.24	105	14699	3.216	ug/l	98
95) Naphthalene	16.13	128	4344	1.901	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD053119\
Data File : VD062564.D
Acq On : 31 May 2019 15:26
Operator : VA/AP
Sample : K3125-21
Misc : 5.04g/5ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
13903

Quant Time: Jun 01 00:11:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D053019S.M
Quant Title : SW846 8260
QLast Update : Fri May 31 02:08:29 2019
Response via : Initial Calibration

