

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD053119\
 Data File : VD062571.D
 Acq On : 31 May 2019 18:36
 Operator : VA/AP
 Sample : K3097-11RE
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20190431-HR-11RE

Quant Time: May 31 23:28:28 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D053019S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 31 02:08:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.24	114	226	50.00	ug/l	0.00
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-12.38
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-14.53
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	445	138.01	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	276.02%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	
62) 4-Bromofluorobenzene	0.00	95	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.75	50	225	75.784	ug/l #	43
16) Acetone	3.31	43	209	329.892	ug/l #	60
23) Vinyl Acetate	5.18	43	196	47.087	ug/l #	80
25) 2-Butanone	5.96	43	210	351.955	ug/l #	62
51) 4-Methyl-2-Pentanone	10.23	43	210	254.983	ug/l #	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD053119\
Data File : VD062571.D
Acq On : 31 May 2019 18:36
Operator : VA/AP
Sample : K3097-11RE
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20190431-HR-11RE

Quant Time: May 31 23:28:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D053019S.M
Quant Title : SW846 8260
QLast Update : Fri May 31 02:08:29 2019
Response via : Initial Calibration

