



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053122\
 Data File : VD073422.D
 Acq On : 31 May 2022 10:10
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Title : SW846 8260
 Last Update : Thu May 19 15:13:42 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1807

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	46448	PASS
75	95	30	60	49.8	126413	PASS
95	95	100	100	100.0	253717	PASS
96	95	5	9	7.2	18165	PASS
173	174	0.00	2	1.2	2391	PASS
174	95	50	100	77.6	196971	PASS
175	174	5	9	7.3	14384	PASS
176	174	95	101	95.4	187904	PASS
177	176	5	9	6.5	12195	PASS