

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053123\
 Data File : VD076295.D
 Acq On : 31 May 2023 20:24
 Operator : KP/SY
 Sample : 02983-05
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB24A

Quant Time: Jun 01 01:44:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	5581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	10443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	9429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	4068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	5997	111.181	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	222.360%#
35) Dibromofluoromethane	7.799	113	6354	88.025	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	176.060%#
50) Toluene-d8	10.269	98	11175	39.713	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.420%
62) 4-Bromofluorobenzene	12.575	95	3789	48.473	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.940%
Target Compounds						
						Qvalue
5) Bromomethane	2.699	94	715	9.783	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053123\
Data File : VD076295.D
Acq On : 31 May 2023 20:24
Operator : KP/SY
Sample : 02983-05
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB24A

Quant Time: Jun 01 01:44:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

