

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053123\
 Data File : VD076296.D
 Acq On : 31 May 2023 20:54
 Operator : KP/SY
 Sample : 02983-06
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB24B

Quant Time: Jun 01 01:44:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	13211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	30764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	34621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	15555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	18306	143.373	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 286.740%#	
35) Dibromofluoromethane	7.804	113	17394	81.798	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 163.600%#	
50) Toluene-d8	10.269	98	30527	36.826	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 73.660%	
62) 4-Bromofluorobenzene	12.575	95	14276	61.997	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 124.000%	
Target Compounds						
					Qvalue	
16) Acetone	4.028	43	8570	337.352	ug/l #	76
25) 2-Butanone	7.087	43	1710	53.535	ug/l #	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053123\
Data File : VD076296.D
Acq On : 31 May 2023 20:54
Operator : KP/SY
Sample : 02983-06
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB24B

Quant Time: Jun 01 01:44:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

