

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
 Data File : VD069283.D
 Acq On : 01 Jun 2021 15:05
 Operator : VA/SY
 Sample : VD0601SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0601SBL01

Quant Time: Jun 02 02:25:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	428088	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	748327	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	676600	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	295864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	190858	46.46	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.92%		
35) Dibromofluoromethane	7.920	113	205009	43.63	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	87.26%		
50) Toluene-d8	10.338	98	700311	40.98	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	81.96%		
62) 4-Bromofluorobenzene	12.626	95	221785	39.61	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	79.22%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	4557	1.01	ug/l	100
16) Acetone	4.156	43	8045	11.22	ug/l #	82
95) Naphthalene	15.396	128	7690	4.01	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
Data File : VD069283.D
Acq On : 01 Jun 2021 15:05
Operator : VA/SY
Sample : VD0601SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0601SBL01

Quant Time: Jun 02 02:25:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

