

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
 Data File : VD069289.D
 Acq On : 01 Jun 2021 17:57
 Operator : VA/SY
 Sample : M2565-01
 Misc : 6.57G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-052821

Quant Time: Jun 02 02:49:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	293350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	573224	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	556545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	248271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	194174	68.98	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.96%
35) Dibromofluoromethane	7.914	113	208212	57.84	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.68%
50) Toluene-d8	10.337	98	706136	53.95	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.90%
62) 4-Bromofluorobenzene	12.631	95	231713	54.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.04%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
Data File : VD069289.D
Acq On : 01 Jun 2021 17:57
Operator : VA/SY
Sample : M2565-01
Misc : 6.57G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-052821

Quant Time: Jun 02 02:49:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

