

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
 Data File : VD069282.D
 Acq On : 01 Jun 2021 14:37
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060121

Quant Time: Jun 02 02:25:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.526	0.506	3.8	99	0.00
3 P	Chloromethane	0.397	0.391	1.5	107	0.00
4 C	Vinyl Chloride	0.405	0.402	0.7#	104	0.00
5 T	Bromomethane	0.295	0.305	-3.4	112	0.00
6 T	Chloroethane	0.231	0.227	1.7	107	0.00
7 T	Trichlorofluoromethane	0.942	0.918	2.5	101	0.00
8 T	Diethyl Ether	0.231	0.251	-8.7	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.554	0.549	0.9	105	0.00
10 T	Methyl Iodide	0.503	0.543	-8.0	97	0.00
11 T	Tert butyl alcohol	0.038	0.031	18.4	123	0.00
12 CM	1,1-Dichloroethene	0.501	0.521	-4.0#	106	0.00
13 T	Acrolein	0.028	0.034	-21.4	108	0.00
14 T	Allyl chloride	0.605	0.642	-6.1	108	0.00
15 T	Acrylonitrile	0.090	0.095	-5.6	107	0.00
16 T	Acetone	0.084	0.073	13.1	96	0.00
17 T	Carbon Disulfide	1.658	1.694	-2.2	103	0.00
18 T	Methyl Acetate	0.227	0.216	4.8	104	0.00
19 T	Methyl tert-butyl Ether	0.921	1.007	-9.3	110	0.00
20 T	Methylene Chloride	0.778	0.656	15.7	111	0.00
21 T	trans-1,2-Dichloroethene	0.594	0.610	-2.7	105	0.00
22 T	Diisopropyl ether	1.162	1.278	-10.0	108	0.00
23 T	Vinyl Acetate	0.669	0.774	-15.7	108	0.00
24 P	1,1-Dichloroethane	0.972	0.988	-1.6	107	0.00
25 T	2-Butanone	0.110	0.115	-4.5	108	0.00
26 T	2,2-Dichloropropane	0.896	0.878	2.0	105	0.00
27 T	cis-1,2-Dichloroethene	0.629	0.670	-6.5	109	0.00
28 T	Bromochloromethane	0.337	0.324	3.9	107	0.00
29 T	Tetrahydrofuran	0.063	0.068	-7.9	107	0.00
30 C	Chloroform	1.063	1.067	-0.4#	106	0.00
31 T	Cyclohexane	0.762	0.751	1.4	103	0.00
32 T	1,1,1-Trichloroethane	0.959	0.960	-0.1	106	0.00
33 S	1,2-Dichloroethane-d4	0.480	0.431	10.2	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
35 S	Dibromofluoromethane	0.314	0.282	10.2	108	0.00
36 T	1,1-Dichloropropene	0.486	0.498	-2.5	104	0.00
37 T	Ethyl Acetate	0.165	0.168	-1.8	106	0.00
38 T	Carbon Tetrachloride	0.525	0.523	0.4	104	0.00
39 T	Methylcyclohexane	0.497	0.560	-12.7	109	0.00
40 TM	Benzene	1.412	1.421	-0.6	106	0.00
41 T	Methacrylonitrile	0.080	0.090	-12.5	124	0.00
42 TM	1,2-Dichloroethane	0.377	0.370	1.9	104	0.00
43 T	Isopropyl Acetate	0.300	0.307	-2.3	105	0.00
44 TM	Trichloroethene	0.384	0.397	-3.4	108	0.00
45 C	1,2-Dichloropropane	0.336	0.343	-2.1#	107	0.00
46 T	Dibromomethane	0.192	0.194	-1.0	103	0.00
47 T	Bromodichloromethane	0.495	0.496	-0.2	106	0.00
48 T	Methyl methacrylate	0.139	0.143	-2.9	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
 Data File : VD069282.D
 Acq On : 01 Jun 2021 14:37
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVV060121

Quant Time: Jun 02 02:25:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	113	0.00
50 S	Toluene-d8	1.142	1.035	9.4	107	0.00
51 T	4-Methyl-2-Pentanone	0.155	0.164	-5.8	106	0.00
52 CM	Toluene	0.891	0.942	-5.7#	107	0.00
53 T	t-1,3-Dichloropropene	0.442	0.453	-2.5	105	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.534	-1.3	106	0.00
55 T	1,1,2-Trichloroethane	0.266	0.262	1.5	104	0.00
56 T	Ethyl methacrylate	0.274	0.303	-10.6	109	0.00
57 T	1,3-Dichloropropane	0.432	0.447	-3.5	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.134	-7.2	120	0.00
59 T	2-Hexanone	0.104	0.111	-6.7	106	0.00
60 T	Dibromochloromethane	0.336	0.338	-0.6	107	0.00
61 T	1,2-Dibromoethane	0.254	0.259	-2.0	108	0.00
62 S	4-Bromofluorobenzene	0.374	0.344	8.0	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	112	0.00
64 T	Tetrachloroethene	0.348	0.344	1.1	105	0.00
65 PM	Chlorobenzene	1.022	1.047	-2.4	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.383	0.385	-0.5	106	0.00
67 C	Ethyl Benzene	1.755	1.895	-8.0#	107	0.00
68 T	m/p-Xylenes	0.688	0.740	-7.6	106	0.00
69 T	o-Xylene	0.608	0.672	-10.5	107	0.00
70 T	Styrene	1.056	1.171	-10.9	106	0.00
71 P	Bromoform	0.199	0.205	-3.0	109	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.422	3.870	-13.1	108	0.00
74 T	N-amyl acetate	0.604	0.658	-8.9	106	0.00
75 P	1,1,2,2-Tetrachloroethane	0.639	0.648	-1.4	103	0.00
76 T	1,2,3-Trichloropropane	0.434	0.411	5.3	102	0.00
77 T	Bromobenzene	0.824	0.871	-5.7	105	0.00
78 T	n-propylbenzene	4.197	4.709	-12.2	105	0.00
79 T	2-Chlorotoluene	2.386	2.623	-9.9	106	0.00
80 T	1,3,5-Trimethylbenzene	2.898	3.262	-12.6	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.196	0.208	-6.1	102	0.00
82 T	4-Chlorotoluene	2.510	2.703	-7.7	105	0.00
83 T	tert-Butylbenzene	2.427	2.778	-14.5	107	0.00
84 T	1,2,4-Trimethylbenzene	2.882	3.276	-13.7	107	0.00
85 T	sec-Butylbenzene	3.582	4.051	-13.1	105	0.00
86 T	p-Isopropyltoluene	3.132	3.567	-13.9	106	0.00
87 T	1,3-Dichlorobenzene	1.668	1.743	-4.5	105	0.00
88 T	1,4-Dichlorobenzene	1.662	1.697	-2.1	103	0.00
89 T	n-Butylbenzene	2.931	3.293	-12.4	104	0.00
90 T	Hexachloroethane	0.595	0.623	-4.7	105	0.00
91 T	1,2-Dichlorobenzene	1.437	1.506	-4.8	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.101	1.0	106	0.00
93 T	1,2,4-Trichlorobenzene	0.862	0.953	-10.6	107	0.00
94 T	Hexachlorobutadiene	0.541	0.571	-5.5	105	0.00
95 T	Naphthalene	1.520	1.742	-14.6	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
Data File : VD069282.D
Acq On : 01 Jun 2021 14:37
Operator : VA/SY
Sample : VSTDICV050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060121

Quant Time: Jun 02 02:25:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.759	0.830	-9.4	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6