



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060121\
 Data File : VD069274.D
 Acq On : 01 Jun 2021 09:50
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Title : SW846 8260
 Last Update : Wed Jun 02 02:17:36 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	15642	PASS
75	95	30	60	52.1	50901	PASS
95	95	100	100	100.0	97712	PASS
96	95	5	9	7.2	7033	PASS
173	174	0.00	2	1.5	1311	PASS
174	95	50	100	88.0	86019	PASS
175	174	5	9	7.6	6570	PASS
176	174	95	101	95.0	81683	PASS
177	176	5	9	7.3	5967	PASS