

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060122\
 Data File : VD073453.D
 Acq On : 01 Jun 2022 17:45
 Operator : VA/SY
 Sample : N3109-03
 Misc : 4.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BB-9

Quant Time: Jun 02 08:52:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	245846	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	449841	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	498792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	240069	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	141700	52.154	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.300%
35) Dibromofluoromethane	7.908	113	157804	53.162	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.320%
50) Toluene-d8	10.332	98	564685	51.607	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.220%
62) 4-Bromofluorobenzene	12.620	95	220851	57.513	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	115.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060122\
Data File : VD073453.D
Acq On : 01 Jun 2022 17:45
Operator : VA/SY
Sample : N3109-03
Misc : 4.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BB-9

Quant Time: Jun 02 08:52:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

