

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060122\
 Data File : VD073455.D
 Acq On : 01 Jun 2022 18:43
 Operator : VA/SY
 Sample : N3108-01
 Misc : 5.25G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NP-1

Quant Time: Jun 02 08:52:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	12949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	23629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	27250	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	13276	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	12181	85.119	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	170.240%#
35) Dibromofluoromethane	7.909	113	10824	69.420	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	138.840%
50) Toluene-d8	10.332	98	28518	49.617	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.620	95	11719	58.099	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060122\
Data File : VD073455.D
Acq On : 01 Jun 2022 18:43
Operator : VA/SY
Sample : N3108-01
Misc : 5.25G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NP-1

Quant Time: Jun 02 08:52:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

