

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060122\
 Data File : VD073458.D
 Acq On : 01 Jun 2022 20:10
 Operator : VA/SY
 Sample : N3104-04
 Misc : 4.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-7

Quant Time: Jun 02 08:53:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	11271	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	18814	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	22855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	10289	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	9542	76.605	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.220%
35) Dibromofluoromethane	7.903	113	8581	69.120	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	138.240%
50) Toluene-d8	10.326	98	23389	51.108	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.220%
62) 4-Bromofluorobenzene	12.620	95	9268	57.707	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	115.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060122\
Data File : VD073458.D
Acq On : 01 Jun 2022 20:10
Operator : VA/SY
Sample : N3104-04
Misc : 4.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-7

Quant Time: Jun 02 08:53:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

