

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060123\
 Data File : VD076304.D
 Acq On : 01 Jun 2023 18:37
 Operator : KP/SY
 Sample : VSTDIC020
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023

Quant Time: Jun 02 05:36:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 05:32:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	171674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	316969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	303393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	162501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	37678	22.520	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	45.040%#		
35) Dibromofluoromethane	7.799	113	42491	20.308	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	40.620%#		
50) Toluene-d8	10.269	98	159605	19.784	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	39.560%#		
62) 4-Bromofluorobenzene	12.575	95	46464	20.649	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	41.300%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.934	85	29732	19.015	ug/l	98
3) Chloromethane	2.146	50	83672	22.658	ug/l	99
4) Vinyl Chloride	2.275	62	84141	20.394	ug/l	97
5) Bromomethane	2.664	94	40152	18.183	ug/l	100
6) Chloroethane	2.816	64	57616	21.045	ug/l	98
7) Trichlorofluoromethane	3.158	101	54893	19.395	ug/l	96
8) Diethyl Ether	3.593	74	22061	21.969	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.952	101	34718	18.852	ug/l	98
10) Methyl Iodide	4.152	142	43646	20.273	ug/l	100
11) Tert butyl alcohol	5.063	59	12196	114.505	ug/l	97
12) 1,1-Dichloroethene	3.928	96	38780	20.279	ug/l	97
13) Acrolein	3.793	56	16049	140.667	ug/l	95
14) Allyl chloride	4.546	41	41238	21.765	ug/l	97
15) Acrylonitrile	5.258	53	47863	115.245	ug/l	99
16) Acetone	4.022	43	32510	115.895	ug/l #	86
17) Carbon Disulfide	4.258	76	129069	20.271	ug/l	98
18) Methyl Acetate	4.558	43	32814	23.286	ug/l	98
19) Methyl tert-butyl Ether	5.316	73	83527	21.645	ug/l	98
20) Methylene Chloride	4.793	84	64170	24.063	ug/l	98
21) trans-1,2-Dichloroethene	5.305	96	47032	21.166	ug/l	97
22) Diisopropyl ether	6.210	45	102346	22.759	ug/l	96
23) Vinyl Acetate	6.152	43	233305	110.134	ug/l	98
24) 1,1-Dichloroethane	6.105	63	77208	21.912	ug/l	97
25) 2-Butanone	7.087	43	51352	117.628	ug/l	98
26) 2,2-Dichloropropane	7.069	77	58248	20.921	ug/l	95
27) cis-1,2-Dichloroethene	7.075	96	52164	21.424	ug/l	98
28) Bromochloromethane	7.422	49	31760	23.427	ug/l	98
29) Tetrahydrofuran	7.440	42	31029	115.669	ug/l	99
30) Chloroform	7.593	83	82532	22.264	ug/l	99
31) Cyclohexane	7.869	56	47798	18.560	ug/l	95
32) 1,1,1-Trichloroethane	7.787	97	59990	20.710	ug/l	98
36) 1,1-Dichloropropene	8.004	75	54207	18.734	ug/l	99
37) Ethyl Acetate	7.169	43	21282	19.397	ug/l	97
38) Carbon Tetrachloride	7.987	117	45264	18.165	ug/l	94
39) Methylcyclohexane	9.275	83	51358	15.753	ug/l	95
40) Benzene	8.246	78	183849	20.019	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060123\
 Data File : VD076304.D
 Acq On : 01 Jun 2023 18:37
 Operator : KP/SY
 Sample : VSTDICC020
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023

Quant Time: Jun 02 05:36:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 05:32:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.404	41	11925m	21.544	ug/l	
42) 1,2-Dichloroethane	8.322	62	46527	20.846	ug/l	98
43) Isopropyl Acetate	8.357	43	40333	20.416	ug/l	100
44) Trichloroethene	9.028	130	47635	18.770	ug/l	98
45) 1,2-Dichloropropane	9.298	63	45354	20.459	ug/l	99
46) Dibromomethane	9.393	93	27783	21.356	ug/l	98
47) Bromodichloromethane	9.581	83	62932	20.648	ug/l	97
48) Methyl methacrylate	9.381	41	16581	19.549	ug/l	96
49) 1,4-Dioxane	9.381	88	6667	423.274	ug/l	95
51) 4-Methyl-2-Pentanone	10.157	43	110233	105.468	ug/l	98
52) Toluene	10.334	92	114111	20.303	ug/l	97
53) t-1,3-Dichloropropene	10.551	75	58434	20.503	ug/l	100
54) cis-1,3-Dichloropropene	10.016	75	67974	20.190	ug/l	98
55) 1,1,2-Trichloroethane	10.734	97	39454	21.304	ug/l	97
56) Ethyl methacrylate	10.598	69	38764	20.046	ug/l	99
57) 1,3-Dichloropropane	10.881	76	61249	20.970	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.869	63	67965	85.286	ug/l	99
59) 2-Hexanone	10.922	43	72071	103.602	ug/l	99
60) Dibromochloromethane	11.069	129	43769	20.623	ug/l	95
61) 1,2-Dibromoethane	11.181	107	35030	20.472	ug/l	99
64) Tetrachloroethene	10.810	164	38942	18.016	ug/l	94
65) Chlorobenzene	11.604	112	130337	19.983	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.681	131	44597	19.762	ug/l	98
67) Ethyl Benzene	11.687	91	194072	18.624	ug/l	97
68) m/p-Xylenes	11.792	106	166314	38.581	ug/l	97
69) o-Xylene	12.122	106	74598	18.862	ug/l	97
70) Styrene	12.134	104	139191	20.392	ug/l	100
71) Bromoform	12.298	173	26916	21.055	ug/l #	99
73) Isopropylbenzene	12.422	105	175207	17.701	ug/l	99
74) N-amyl acetate	12.234	43	36965	19.111	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.675	83	45807	20.289	ug/l	98
76) 1,2,3-Trichloropropane	12.722	75	30312m	20.680	ug/l	
77) Bromobenzene	12.698	156	52165	19.063	ug/l	95
78) n-propylbenzene	12.763	91	224899	18.202	ug/l	99
79) 2-Chlorotoluene	12.845	91	130263	18.900	ug/l	99
80) 1,3,5-Trimethylbenzene	12.904	105	153612	18.645	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.469	75	12774	20.044	ug/l	97
82) 4-Chlorotoluene	12.945	91	143396	19.406	ug/l	99
83) tert-Butylbenzene	13.169	119	124045	17.722	ug/l	97
84) 1,2,4-Trimethylbenzene	13.210	105	158430	18.840	ug/l	99
85) sec-Butylbenzene	13.345	105	188210	17.693	ug/l	100
86) p-Isopropyltoluene	13.463	119	160381	18.029	ug/l	99
87) 1,3-Dichlorobenzene	13.457	146	108491	19.586	ug/l	99
88) 1,4-Dichlorobenzene	13.539	146	111734	19.863	ug/l	99
89) n-Butylbenzene	13.786	91	141202	17.282	ug/l	99
90) Hexachloroethane	14.051	117	34114	18.963	ug/l	93
91) 1,2-Dichlorobenzene	13.834	146	97176	19.940	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.451	75	6115	20.946	ug/l	98
93) 1,2,4-Trichlorobenzene	15.098	180	51370	18.419	ug/l	97
94) Hexachlorobutadiene	15.204	225	22129	16.929	ug/l	96
95) Naphthalene	15.333	128	100459	18.734	ug/l	99
96) 1,2,3-Trichlorobenzene	15.522	180	47321	18.482	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060123\
 Data File : VD076304.D
 Acq On : 01 Jun 2023 18:37
 Operator : KP/SY
 Sample : VSTDICC020
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
ClientSampleId :
 VSTDICC020


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023

Quant Time: Jun 02 05:36:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 05:32:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.869 min Scan# 1068
Delta R.T. 0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 171674
Ion Ratio Lower Upper
168 100
99 48.6 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#2
Dichlorodifluoromethane
Concen: 19.015 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 85 Resp: 29732
Ion Ratio Lower Upper
85 100
87 32.1 15.7 47.0





#3
Chloromethane
Concen: 22.658 ug/l
RT: 2.146 min Scan# 95
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 50 Resp: 83672
Ion Ratio Lower Upper
50 100
52 34.0 26.6 40.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#4
Vinyl Chloride
Concen: 20.394 ug/l
RT: 2.275 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Tgt Ion: 62 Resp: 84141
Ion Ratio Lower Upper
62 100
64 30.7 26.0 39.0





#5
Bromomethane
Concen: 18.183 ug/l
RT: 2.664 min Scan# 11
Delta R.T. -0.012 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 94 Resp: 40152
Ion Ratio Lower Upper
94 100
96 95.0 75.8 113.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#6
Chloroethane
Concen: 21.045 ug/l
RT: 2.816 min Scan# 209
Delta R.T. -0.007 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion: 64 Resp: 57616
Ion Ratio Lower Upper
64 100
66 31.0 25.5 38.3









#7
 Trichlorofluoromethane
 Concen: 19.395 ug/l
 RT: 3.158 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020

Tgt Ion:101 Resp: 5489
 Ion Ratio Lower Upper
 101 100
 103 62.1 52.3 78.5




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023



#8
 Diethyl Ether
 Concen: 21.969 ug/l
 RT: 3.593 min Scan# 341
 Delta R.T. 0.012 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37

Tgt Ion: 74 Resp: 22061
 Ion Ratio Lower Upper
 74 100
 45 78.0 37.5 112.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.852 ug/l
RT: 3.952 min Scan# 402
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion:101 Resp: 34718
Ion Ratio Lower Upper
101 100
85 42.5 33.8 50.6
151 72.5 59.8 89.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#10
Methyl Iodide
Concen: 20.273 ug/l
RT: 4.152 min Scan# 436
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:142 Resp: 43646
Ion Ratio Lower Upper
142 100
127 36.8 29.4 44.0
141 13.9 11.1 16.7









#11
Tert butyl alcohol
Concen: 114.505 ug/l
RT: 5.063 min Scan# 591
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 59 Resp: 12190
Ion Ratio Lower Upper
59 100
57 8.7 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#12
1,1-Dichloroethene
Concen: 20.279 ug/l
RT: 3.928 min Scan# 398
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 96 Resp: 38780
Ion Ratio Lower Upper
96 100
61 127.8 100.6 150.8
98 59.2 50.2 75.2





#13
Acrolein
Concen: 140.667 ug/l
RT: 3.793 min Scan# 31
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion: 56 Resp: 16049
Ion Ratio Lower Upper
56 100
55 72.9 55.1 82.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#14
Allyl chloride
Concen: 21.765 ug/l
RT: 4.546 min Scan# 503
Delta R.T. -0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion: 41 Resp: 41238
Ion Ratio Lower Upper
41 100
39 63.0 53.8 80.6
76 57.7 45.8 68.8









#15
Acrylonitrile
Concen: 115.245 ug/l
RT: 5.258 min Scan# 61
Delta R.T. 0.012 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 53 Resp: 4786
Ion Ratio Lower Upper
53 100
52 80.5 65.0 97.4
51 33.6 28.3 42.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#16
Acetone
Concen: 115.895 ug/l
RT: 4.022 min Scan# 414
Delta R.T. -0.001 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 43 Resp: 32510
Ion Ratio Lower Upper
43 100
58 34.5 34.7 52.1#





#17
Carbon Disulfide
Concen: 20.271 ug/l
RT: 4.258 min Scan# 454
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 76 Resp: 129069
Ion Ratio Lower Upper
76 100
78 9.1 6.8 10.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#18
Methyl Acetate
Concen: 23.286 ug/l
RT: 4.558 min Scan# 505
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion: 43 Resp: 32814
Ion Ratio Lower Upper
43 100
74 32.1 26.4 39.6





#19

Methyl tert-butyl Ether

Concen: 21.645 ug/l

RT: 5.316 min Scan# 61

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion: 73 Resp: 8352

Ion Ratio Lower Upper

73 100





57 20.1 16.6 25.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#20

Methylene Chloride

Concen: 24.063 ug/l

RT: 4.793 min Scan# 545

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Tgt Ion: 84 Resp: 64170




Ion Ratio Lower Upper


84 100

49 92.8 71.4 107.2

51 29.7 24.5 36.7

86 61.1 48.7 73.1









#21
trans-1,2-Dichloroethene
Concen: 21.166 ug/l
RT: 5.305 min Scan# 611
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 96 Resp: 47032
Ion Ratio Lower Upper
96 100
61 111.9 89.3 133.9
98 67.0 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#22
Diisopropyl ether
Concen: 22.759 ug/l
RT: 6.210 min Scan# 786
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 45 Resp: 102346
Ion Ratio Lower Upper
45 100
43 52.0 40.0 60.0
87 30.7 28.2 42.4
59 13.1 11.0 16.4





#23
 Vinyl Acetate
 Concen: 110.134 ug/l
 RT: 6.152 min Scan# 776
 Delta R.T. 0.006 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020




Tgt Ion: 43 Resp: 23330
 Ion Ratio Lower Upper
 43 100
 86 14.8 12.5 18.7

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023




#24
 1,1-Dichloroethane
 Concen: 21.912 ug/l
 RT: 6.105 min Scan# 768
 Delta R.T. 0.006 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37



Tgt Ion: 63 Resp: 77208
 Ion Ratio Lower Upper
 63 100
 98 10.3 4.5 13.7
 100 4.1 2.5 7.5









#25
2-Butanone
Concen: 117.628 ug/l
RT: 7.087 min Scan# 91
Delta R.T. 0.012 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 5135
Ion Ratio Lower Upper
43 100
72 37.5 30.7 46.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#26
2,2-Dichloropropane
Concen: 20.921 ug/l
RT: 7.069 min Scan# 932
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 77 Resp: 58248
Ion Ratio Lower Upper
77 100
97 24.1 13.2 39.6





#27
 cis-1,2-Dichloroethene
 Concen: 21.424 ug/l
 RT: 7.075 min Scan# 911
 Delta R.T. 0.006 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICC020



Tgt Ion: 96 Resp: 52164
 Ion Ratio Lower Upper
 96 100
 61 121.4 0.0 239.2
 98 65.5 0.0 128.6




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023



#28
 Bromochloromethane
 Concen: 23.427 ug/l
 RT: 7.422 min Scan# 992
 Delta R.T. -0.000 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37

Tgt Ion: 49 Resp: 31760
 Ion Ratio Lower Upper
 49 100
 129 2.5 0.0 6.0
 130 106.3 83.2 124.8





#29

Tetrahydrofuran

Concen: 115.669 ug/l

RT: 7.440 min Scan# 995

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion: 42 Resp: 31029





Ion	Ratio	Lower	Upper
42	100		
72	58.0	46.5	69.7
71	54.5	44.5	66.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#30

Chloroform

Concen: 22.264 ug/l

RT: 7.593 min Scan# 1021




Delta R.T. -0.000 min


Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Tgt Ion: 83 Resp: 82532


Ion	Ratio	Lower	Upper
83	100		
85	66.4	53.4	80.2





#31
Cyclohexane
Concen: 18.560 ug/l
RT: 7.869 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion: 56 Resp: 47798
Ion Ratio Lower Upper
56 100
69 39.5 28.6 43.0
84 110.6 84.8 127.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#32
1,1,1-Trichloroethane
Concen: 20.710 ug/l
RT: 7.787 min Scan# 1054
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion: 97 Resp: 59990
Ion Ratio Lower Upper
97 100
99 64.0 52.0 78.0
61 40.2 33.8 50.8





#33

1,2-Dichloroethane-d4

Concen: 22.520 ug/l

RT: 8.228 min Scan# 1128

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 37678

Ion Ratio Lower Upper

65 100





67 57.8 0.0 115.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37




Tgt Ion:114 Resp: 316969


Ion Ratio Lower Upper

114 100

63 16.9 0.0 33.4

88 13.4 0.0 26.2





#35

Dibromofluoromethane

Concen: 20.308 ug/l

RT: 7.799 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 113 Resp: 4249





Ion	Ratio	Lower	Upper
113	100		
111	101.1	84.7	127.1
192	21.2	16.8	25.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#36

1,1-Dichloropropene

Concen: 18.734 ug/l

RT: 8.004 min Scan# 1091




Delta R.T. 0.006 min


Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Tgt Ion: 75 Resp: 54207


Ion	Ratio	Lower	Upper
75	100		
110	38.7	19.9	59.6
77	30.8	24.7	37.1





#37
Ethyl Acetate
Concen: 19.397 ug/l
RT: 7.169 min Scan# 948
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 2128
Ion Ratio Lower Upper
43 100
61 19.1 14.9 22.3
70 15.0 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#38
Carbon Tetrachloride
Concen: 18.165 ug/l
RT: 7.987 min Scan# 1088
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:117 Resp: 45264
Ion Ratio Lower Upper
117 100
119 91.5 78.0 117.0
121 33.5 24.2 36.2





#39

Methylcyclohexane

Concen: 15.753 ug/l

RT: 9.275 min Scan# 1137

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020





Tgt Ion: 83	Resp: 5135
Ion Ratio	Lower Upper
83 100	
55 67.4	49.4 74.0
98 49.2	38.4 57.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#40

Benzene

Concen: 20.019 ug/l




RT: 8.246 min Scan# 1132


Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37


Tgt Ion: 78	Resp: 183849
Ion Ratio	Lower Upper
78 100	
77 27.1	19.4 29.0





#41
Methacrylonitrile
Concen: 21.544 ug/l m
RT: 7.404 min Scan# 989
Delta R.T. 0.012 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion: 41 Resp: 1192
Ion Ratio Lower Upper
41 100
39 50.2 34.0 51.0
67 102.1 77.0 115.4
52 35.6 27.9 41.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#42
1,2-Dichloroethane
Concen: 20.846 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 62 Resp: 46527
Ion Ratio Lower Upper
62 100
98 11.9 0.0 25.4





#43
Isopropyl Acetate
Concen: 20.416 ug/l
RT: 8.357 min Scan# 1151
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 43 Resp: 4033
Ion Ratio Lower Upper
43 100
61 39.3 31.4 47.0
87 19.0 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#44
Trichloroethene
Concen: 18.770 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion:130 Resp: 47635
Ion Ratio Lower Upper
130 100
95 91.2 0.0 179.2





#45

1,2-Dichloropropane

Concen: 20.459 ug/l

RT: 9.298 min Scan# 1311

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 63 Resp: 45354

Ion Ratio Lower Upper

63 100





65 31.0 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#46

Dibromomethane

Concen: 21.356 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37




Tgt Ion: 93 Resp: 27783


Ion Ratio Lower Upper

93 100

95 83.7 67.7 101.5


174 103.4 85.4 128.2





#47
Bromodichloromethane
Concen: 20.648 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 83 Resp: 6293
Ion Ratio Lower Upper
83 100
85 60.1 50.2 75.4
127 8.9 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#48
Methyl methacrylate
Concen: 19.549 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion: 41 Resp: 16581
Ion Ratio Lower Upper
41 100
69 131.3 102.2 153.2
39 57.2 41.4 62.2





#49

1,4-Dioxane

Concen: 423.274 ug/l

RT: 9.381 min Scan# 11

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020





Tgt Ion:	88	Resp:	666
Ion Ratio	Lower	Upper	
88	100		
43	21.3	19.0	28.6
58	58.1	43.5	65.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#50

Toluene-d8

Concen: 19.784 ug/l




RT: 10.269 min Scan# 1476


Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Tgt Ion:	98	Resp:	159605
Ion Ratio	Lower	Upper	
98	100		
100	64.2	50.3	75.5









#51
4-Methyl-2-Pentanone
Concen: 105.468 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 43 Resp: 110233
Ion Ratio Lower Upper
43 100
58 47.1 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#52
Toluene
Concen: 20.303 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion: 92 Resp: 114111
Ion Ratio Lower Upper
92 100
91 174.1 143.0 214.4









#53
t-1,3-Dichloropropene
Concen: 20.503 ug/l
RT: 10.551 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 75 Resp: 58434
Ion Ratio Lower Upper
75 100
77 31.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#54
cis-1,3-Dichloropropene
Concen: 20.190 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion: 75 Resp: 67974
Ion Ratio Lower Upper
75 100
77 33.2 25.5 38.3
39 29.0 22.1 33.1





#55
1,1,2-Trichloroethane
Concen: 21.304 ug/l
RT: 10.734 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020



Tgt Ion	Ratio	Lower	Upper
97	100		
83	85.2	65.8	98.6
85	57.7	45.7	68.5
99	62.3	46.2	69.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#56
Ethyl methacrylate
Concen: 20.046 ug/l
RT: 10.598 min Scan# 1532
Delta R.T. -0.001 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion	Ratio	Lower	Upper
69	100		
41	49.7	39.1	58.7
39	24.7	20.0	30.0





#57

1,3-Dichloropropane

Concen: 20.970 ug/l

RT: 10.881 min Scan# 1579

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 61249

Ion Ratio Lower Upper

76 100





78 33.3 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#58

2-Chloroethyl Vinyl ether

Concen: 85.286 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. 0.000 min

Lab File: VD076304.D




Acq: 01 Jun 2023 18:37


Tgt Ion: 63 Resp: 67965

Ion Ratio Lower Upper

63 100

106 32.0 25.3 37.9









#59
2-Hexanone
Concen: 103.602 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 43 Resp: 7207
Ion Ratio Lower Upper
43 100
58 67.5 34.1 102.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#60
Dibromochloromethane
Concen: 20.623 ug/l
RT: 11.069 min Scan# 1612
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:129 Resp: 43769
Ion Ratio Lower Upper
129 100
127 78.5 37.1 111.5





#61

1,2-Dibromoethane

Concen: 20.472 ug/l

RT: 11.181 min Scan# 1631

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion:107 Resp: 35030

Ion Ratio Lower Upper

107 100





109 95.9 75.6 113.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#62

4-Bromofluorobenzene

Concen: 20.649 ug/l

RT: 12.575 min Scan# 1868

Delta R.T. 0.006 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37




Tgt Ion: 95 Resp: 46464


Ion Ratio Lower Upper

95 100

174 95.0 0.0 201.8


176 87.9 0.0 191.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:117 Resp: 30339
Ion Ratio Lower Upper
117 100
82 50.9 38.3 57.5
119 33.9 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#64
Tetrachloroethene
Concen: 18.016 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion:164 Resp: 38942
Ion Ratio Lower Upper
164 100
166 129.7 99.2 148.8
129 91.0 68.4 102.6
131 84.9 63.8 95.8









#65
Chlorobenzene
Concen: 19.983 ug/l
RT: 11.604 min Scan# 1112
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:112 Resp: 13033
Ion Ratio Lower Upper
112 100
114 34.3 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 19.762 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:131 Resp: 44597
Ion Ratio Lower Upper
131 100
133 95.3 48.5 145.5
119 72.3 35.6 106.9









#67
Ethyl Benzene
Concen: 18.624 ug/l
RT: 11.687 min Scan# 1716
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 194072
Ion Ratio Lower Upper
91 100
106 34.6 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#68
m/p-Xylenes
Concen: 38.581 ug/l
RT: 11.792 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion:106 Resp: 166314
Ion Ratio Lower Upper
106 100
91 188.1 147.0 220.6





#69
o-Xylene
Concen: 18.862 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion:106 Resp: 74598
Ion Ratio Lower Upper
106 100
91 200.6 97.9 293.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#70
Styrene
Concen: 20.392 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:104 Resp: 139191
Ion Ratio Lower Upper
104 100
78 41.1 33.2 49.8
103 54.5 43.6 65.4









#71
Bromoform
Concen: 21.055 ug/l
RT: 12.298 min Scan# 1821
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:173 Resp: 26910
Ion Ratio Lower Upper
173 100
175 48.5 24.6 73.8
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:152 Resp: 162501
Ion Ratio Lower Upper
152 100
115 50.9 27.1 81.3
150 161.2 0.0 353.6









#73
Isopropylbenzene
Concen: 17.701 ug/l
RT: 12.422 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:105 Resp: 17520
Ion Ratio Lower Upper
105 100
120 28.3 14.3 42.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#74
N-ethyl acetate
Concen: 19.111 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion: 43 Resp: 36965
Ion Ratio Lower Upper
43 100
70 56.6 45.0 67.4
55 29.0 23.8 35.6
61 30.1 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 20.289 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 83 Resp: 4580
Ion Ratio Lower Upper
83 100
131 10.2 5.1 15.3
85 67.3 32.6 97.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#76
1,2,3-Trichloropropane
Concen: 20.680 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion: 75 Resp: 30312
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 19.063 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:156 Resp: 5216
Ion Ratio Lower Upper
156 100
77 162.4 76.8 230.4
158 95.9 48.6 145.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#78
n-propylbenzene
Concen: 18.202 ug/l
RT: 12.763 min Scan# 1900
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion: 91 Resp: 224899
Ion Ratio Lower Upper
91 100
120 24.2 12.4 37.2









#79
2-Chlorotoluene
Concen: 18.900 ug/l
RT: 12.845 min Scan# 1914
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 13026
Ion Ratio Lower Upper
91 100
126 39.0 19.4 58.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#80
1,3,5-Trimethylbenzene
Concen: 18.645 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37


Tgt Ion:105 Resp: 153612
Ion Ratio Lower Upper
105 100
120 54.0 26.4 79.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.044 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 75 Resp: 12774
Ion Ratio Lower Upper
75 100
53 79.3 63.0 94.4
89 42.4 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023




#82
4-Chlorotoluene
Concen: 19.406 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37



Tgt Ion: 91 Resp: 143396
Ion Ratio Lower Upper
91 100
126 38.1 19.4 58.4









#83
tert-Butylbenzene
Concen: 17.722 ug/l
RT: 13.169 min Scan# 1969
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion:119 Resp: 124049
Ion Ratio Lower Upper
119 100
91 55.8 29.3 87.8
134 27.5 14.1 42.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#84
1,2,4-Trimethylbenzene
Concen: 18.840 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:105 Resp: 158430
Ion Ratio Lower Upper
105 100
120 49.4 24.3 72.9









#85
 sec-Butylbenzene
 Concen: 17.693 ug/l
 RT: 13.345 min Scan# 1999
 Delta R.T. -0.000 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020

Tgt Ion:105 Resp: 188210
 Ion Ratio Lower Upper
 105 100
 134 22.0 10.9 32.9




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023



#86
 p-Isopropyltoluene
 Concen: 18.029 ug/l
 RT: 13.463 min Scan# 2019
 Delta R.T. 0.006 min
 Lab File: VD076304.D
 Acq: 01 Jun 2023 18:37

Tgt Ion:119 Resp: 160381
 Ion Ratio Lower Upper
 119 100
 134 27.8 14.2 42.6
 91 20.8 10.5 31.5









#87
1,3-Dichlorobenzene
Concen: 19.586 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion:146 Resp: 10849
Ion Ratio Lower Upper
146 100
111 36.0 17.9 53.8
148 63.5 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#88
1,4-Dichlorobenzene
Concen: 19.863 ug/l
RT: 13.539 min Scan# 2032
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:146 Resp: 111734
Ion Ratio Lower Upper
146 100
111 36.0 17.4 52.3
148 62.8 31.7 95.1





#89

n-Butylbenzene

Concen: 17.282 ug/l

RT: 13.786 min Scan# 2074

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020





Tgt Ion:	91	Resp:	14120
Ion Ratio	Lower	Upper	
91	100		
92	53.7	26.5	79.5
134	27.8	13.8	41.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#90

Hexachloroethane

Concen: 18.963 ug/l




RT: 14.051 min Scan# 2119


Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Tgt Ion:	117	Resp:	34114
Ion Ratio	Lower	Upper	
117	100		
201	70.7	38.5	115.5









#91
1,2-Dichlorobenzene
Concen: 19.940 ug/l
RT: 13.834 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:146 Resp: 97170
Ion Ratio Lower Upper
146 100
111 37.0 18.3 54.8
148 63.1 31.6 94.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.946 ug/l
RT: 14.451 min Scan# 2187
Delta R.T. 0.006 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion: 75 Resp: 6115
Ion Ratio Lower Upper
75 100
155 97.3 47.9 143.8
157 121.6 59.9 179.7





#93

1,2,4-Trichlorobenzene

Concen: 18.419 ug/l

RT: 15.098 min Scan# 2297

Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020





Tgt Ion:180	Resp:	51370
Ion Ratio	Lower	Upper
180	100	
182	94.5	48.5 145.6
145	27.1	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/02/2023

Supervised By :Mahesh Dadoda 06/02/2023



#94

Hexachlorobutadiene

Concen: 16.929 ug/l




RT: 15.204 min Scan# 2315


Delta R.T. -0.000 min

Lab File: VD076304.D

Acq: 01 Jun 2023 18:37

Tgt Ion:225	Resp:	22129
Ion Ratio	Lower	Upper
225	100	
223	60.4	31.6 94.8
227	66.0	31.7 95.1









#95
Naphthalene
Concen: 18.734 ug/l
RT: 15.333 min Scan# 2337
Delta R.T. -0.001 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion:128 Resp: 100459
Ion Ratio Lower Upper
128 100
127 12.9 10.5 15.7
129 11.6 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023



#96
1,2,3-Trichlorobenzene
Concen: 18.482 ug/l
RT: 15.522 min Scan# 2369
Delta R.T. -0.000 min
Lab File: VD076304.D
Acq: 01 Jun 2023 18:37

Tgt Ion:180 Resp: 47321
Ion Ratio Lower Upper
180 100
182 94.1 47.3 141.9
145 28.9 15.1 45.3

