



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060123\
Data File : VD076301.D
Acq On : 01 Jun 2023 10:26
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060123S.M
Title : SW846 8260
Last Update : Fri Jun 02 05:50:45 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.6	7159	PASS
75	95	30	60	50.6	20571	PASS
95	95	100	100	100.0	40675	PASS
96	95	5	9	6.6	2699	PASS
173	174	0.00	2	1.4	574	PASS
174	95	50	100	99.7	40573	PASS
175	174	5	9	7.3	2969	PASS
176	174	95	101	95.4	38693	PASS
177	176	5	9	5.9	2300	PASS