

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069295.D
 Acq On : 02 Jun 2021 12:15
 Operator : VA/SY
 Sample : VD0602SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Jun 03 01:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	431423	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	689261	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	652292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	322925	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	184000	44.45	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.90%
35) Dibromofluoromethane	7.914	113	199503	46.09	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.18%
50) Toluene-d8	10.338	98	728267	46.27	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.54%
62) 4-Bromofluorobenzene	12.626	95	244379	47.38	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.76%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.985	85	84287	18.56	ug/l 97
3) Chloromethane	2.209	50	67395	19.66	ug/l 94
4) Vinyl Chloride	2.350	62	70350	20.14	ug/l 91
5) Bromomethane	2.762	94	52945	20.78	ug/l 95
6) Chloroethane	2.920	64	40619	20.38	ug/l 89
7) Trichlorofluoromethane	3.267	101	160562	19.76	ug/l 94
8) Diethyl Ether	3.709	74	38512	19.30	ug/l 99
9) 1,1,2-Trichlorotrifluo...	4.091	101	94024	19.65	ug/l 98
10) Methyl Iodide	4.303	142	72846	16.68	ug/l 96
11) Tert butyl alcohol	5.203	59	37239	134.66	ug/l # 88
12) 1,1-Dichloroethene	4.073	96	83297	19.25	ug/l 94
13) Acrolein	3.920	56	20229	85.32	ug/l 100
14) Allyl chloride	4.720	41	95895	18.37	ug/l 99
15) Acrylonitrile	5.414	53	72859	93.64	ug/l 98
16) Acetone	4.156	43	70527	97.63	ug/l # 85
17) Carbon Disulfide	4.409	76	266014	18.59	ug/l 100
18) Methyl Acetate	4.714	43	36584	18.71	ug/l # 72
19) Methyl tert-butyl Ether	5.473	73	146329	18.41	ug/l # 91
20) Methylene Chloride	4.967	84	127177	20.00	ug/l 95
21) trans-1,2-Dichloroethene	5.473	96	98012	19.12	ug/l 95
22) Diisopropyl ether	6.361	45	194273	19.37	ug/l # 93
23) Vinyl Acetate	6.303	43	525080	89.76	ug/l 99
24) 1,1-Dichloroethane	6.261	63	161055	19.19	ug/l 98
25) 2-Butanone	7.208	43	90293	94.99	ug/l 96
26) 2,2-Dichloropropane	7.208	77	146541	18.95	ug/l 97
27) cis-1,2-Dichloroethene	7.208	96	103385	19.05	ug/l 98
28) Bromochloromethane	7.544	49	60674	20.86	ug/l 99
29) Tetrahydrofuran	7.561	42	49528	90.65	ug/l 100
30) Chloroform	7.708	83	177574	19.36	ug/l 98
31) Cyclohexane	7.985	56	118672	18.04	ug/l # 93
32) 1,1,1-Trichloroethane	7.903	97	157274	19.00	ug/l 98
36) 1,1-Dichloropropene	8.114	75	132899	19.84	ug/l 100
37) Ethyl Acetate	7.297	43	43678	19.20	ug/l 98
38) Carbon Tetrachloride	8.097	117	143486	19.84	ug/l 96
39) Methylcyclohexane	9.355	83	132684	19.36	ug/l 96
40) Benzene	8.350	78	378394	19.45	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069295.D
 Acq On : 02 Jun 2021 12:15
 Operator : VA/SY
 Sample : VD0602SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Jun 03 01:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.526	41	23210	21.02	ug/l	87
42) 1,2-Dichloroethane	8.426	62	102262	19.69	ug/l	100
43) Isopropyl Acetate	8.450	43	78332	18.92	ug/l	100
44) Trichloroethene	9.114	130	104353	19.74	ug/l	95
45) 1,2-Dichloropropane	9.385	63	92879	20.05	ug/l	99
46) Dibromomethane	9.473	93	51395	19.41	ug/l	95
47) Bromodichloromethane	9.661	83	135274	19.81	ug/l	99
48) Methyl methacrylate	9.461	41	34045	17.71	ug/l	84
49) 1,4-Dioxane	9.461	88	9422	362.24	ug/l #	88
51) 4-Methyl-2-Pentanone	10.226	43	208326	97.42	ug/l	98
52) Toluene	10.402	92	243708	19.84	ug/l	100
53) t-1,3-Dichloropropene	10.620	75	118050	19.36	ug/l	99
54) cis-1,3-Dichloropropene	10.085	75	141469	19.46	ug/l	95
55) 1,1,2-Trichloroethane	10.796	97	73694	20.06	ug/l	98
56) Ethyl methacrylate	10.661	69	69798	18.42	ug/l	99
57) 1,3-Dichloropropane	10.944	76	121265	20.36	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.944	63	188398	109.54	ug/l	100
59) 2-Hexanone	10.985	43	141406	98.93	ug/l	98
60) Dibromochloromethane	11.138	129	90923	19.65	ug/l	98
61) 1,2-Dibromoethane	11.244	107	70329	20.07	ug/l	98
64) Tetrachloroethene	10.873	164	87766	19.33	ug/l	96
65) Chlorobenzene	11.667	112	265170	19.89	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.738	131	96291	19.28	ug/l	98
67) Ethyl Benzene	11.743	91	444750	19.42	ug/l	97
68) m/p-Xylenes	11.855	106	354679	39.52	ug/l	99
69) o-Xylene	12.179	106	156436	19.71	ug/l	98
70) Styrene	12.191	104	272618	19.79	ug/l	99
71) Bromoform	12.355	173	52498	20.27	ug/l #	100
73) Isopropylbenzene	12.473	105	419815	19.00	ug/l	100
74) N-amyl acetate	12.285	43	70553	18.08	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.726	83	79657	19.31	ug/l	98
76) 1,2,3-Trichloropropane	12.779	75	99237	35.37	ug/l #	1
77) Bromobenzene	12.761	156	104308	19.60	ug/l	100
78) n-propylbenzene	12.814	91	527213	19.45	ug/l	99
79) 2-Chlorotoluene	12.902	91	298998	19.41	ug/l	98
80) 1,3,5-Trimethylbenzene	12.955	105	365037	19.50	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.526	75	24426	19.29	ug/l	100
82) 4-Chlorotoluene	13.002	91	319455	19.71	ug/l	100
83) tert-Butylbenzene	13.220	119	299590	19.11	ug/l	98
84) 1,2,4-Trimethylbenzene	13.267	105	369353	19.84	ug/l	99
85) sec-Butylbenzene	13.396	105	451920	19.53	ug/l	99
86) p-Isopropyltoluene	13.508	119	390867	19.32	ug/l	99
87) 1,3-Dichlorobenzene	13.508	146	212596	19.74	ug/l	99
88) 1,4-Dichlorobenzene	13.590	146	208768	19.45	ug/l	97
89) n-Butylbenzene	13.837	91	358488	18.94	ug/l	99
90) Hexachloroethane	14.102	117	73684	19.17	ug/l	99
91) 1,2-Dichlorobenzene	13.885	146	178833	19.27	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.496	75	12240	18.62	ug/l	96
93) 1,2,4-Trichlorobenzene	15.155	180	106641	19.15	ug/l	100
94) Hexachlorobutadiene	15.261	225	66491	19.02	ug/l	99
95) Naphthalene	15.390	128	171973	17.88	ug/l	98
96) 1,2,3-Trichlorobenzene	15.584	180	91189	18.61	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
Data File : VD069295.D
Acq On : 02 Jun 2021 12:15
Operator : VA/SY
Sample : VD0602SBS01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :


Quant Time: Jun 03 01:06:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 03 01:06:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:168 Resp: 431423
Ion Ratio Lower Upper
168 100
99 55.4 46.2 69.2



#2
Dichlorodifluoromethane
Concen: 18.56 ug/l
RT: 1.985 min Scan# 11
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 85 Resp: 84287
Ion Ratio Lower Upper
85 100
87 30.4 15.9 47.7









#3
Chloromethane
Concen: 19.66 ug/l
RT: 2.209 min Scan# 49
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 50 Resp: 67395
Ion Ratio Lower Upper
50 100
52 30.2 26.7 40.1



#4
Vinyl Chloride
Concen: 20.14 ug/l
RT: 2.350 min Scan# 73
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Tgt Ion: 62 Resp: 70350
Ion Ratio Lower Upper
62 100
64 27.7 26.2 39.4








#5
Bromomethane
Concen: 20.78 ug/l
RT: 2.762 min Scan# 143
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion: 94 Resp: 52945
Ion Ratio Lower Upper
94 100
96 92.5 70.3 105.5




#6
Chloroethane
Concen: 20.38 ug/l
RT: 2.920 min Scan# 170
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15



Tgt Ion: 64 Resp: 40619
Ion Ratio Lower Upper
64 100
66 37.8 25.4 38.0









#7
Trichlorofluoromethane
Concen: 19.76 ug/l
RT: 3.267 min Scan# 229
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 160562
Ion Ratio Lower Upper
101 100
103 58.7 51.0 76.4



#8
Diethyl Ether
Concen: 19.30 ug/l
RT: 3.709 min Scan# 304
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 74 Resp: 38512
Ion Ratio Lower Upper
74 100
45 85.5 42.4 127.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.65 ug/l
RT: 4.091 min Scan# 369
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 94024
Ion Ratio Lower Upper
101 100
85 40.4 35.1 52.7
151 75.7 60.2 90.2



#10
Methyl Iodide
Concen: 16.68 ug/l
RT: 4.303 min Scan# 405
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Tgt Ion:142 Resp: 72846
Ion Ratio Lower Upper
142 100
127 42.5 36.2 54.2
141 14.5 12.2 18.2








#11
Tert butyl alcohol
Concen: 134.66 ug/l
RT: 5.203 min Scan# 558
Delta R.T. -0.017 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Instrument :
MSVOA_D
ClientSampled :







Tgt Ion: 59 Resp: 37239
Ion Ratio Lower Upper
59 100
57 5.5 7.8 11.8#



#12
1,1-Dichloroethene
Concen: 19.25 ug/l
RT: 4.073 min Scan# 366
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 96 Resp: 83297
Ion Ratio Lower Upper
96 100
61 138.3 116.5 174.7
98 58.2 50.8 76.2









#13
Acrolein
Concen: 85.32 ug/l
RT: 3.920 min Scan# 340
Delta R.T. 0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 56 Resp: 20229
Ion Ratio Lower Upper
56 100
55 67.7 53.8 80.8



#14
Allyl chloride
Concen: 18.37 ug/l
RT: 4.720 min Scan# 476
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 41 Resp: 95895
Ion Ratio Lower Upper
41 100
39 68.4 56.1 84.1
76 47.7 37.9 56.9









#15
Acrylonitrile
Concen: 93.64 ug/l
RT: 5.414 min Scan# 594
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 53 Resp: 72859
Ion Ratio Lower Upper
53 100
52 81.4 64.5 96.7
51 38.1 28.6 43.0



#16
Acetone
Concen: 97.63 ug/l
RT: 4.156 min Scan# 380
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Tgt Ion: 43 Resp: 70527
Ion Ratio Lower Upper
43 100
58 42.8 27.5 41.3#








#17
Carbon Disulfide
Concen: 18.59 ug/l
RT: 4.409 min Scan# 423
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion: 76 Resp: 266014
Ion Ratio Lower Upper
76 100
78 9.2 7.4 11.0




#18
Methyl Acetate
Concen: 18.71 ug/l
RT: 4.714 min Scan# 475
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15



Tgt Ion: 43 Resp: 36584
Ion Ratio Lower Upper
43 100
74 13.5 22.8 34.2#









#19
Methyl tert-butyl Ether
Concen: 18.41 ug/l
RT: 5.473 min Scan# 604
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 73 Resp: 146329
Ion Ratio Lower Upper
73 100
57 23.8 15.7 23.5#



#20
Methylene Chloride
Concen: 20.00 ug/l
RT: 4.967 min Scan# 518
Delta R.T. 0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 84 Resp: 127177
Ion Ratio Lower Upper
84 100
49 95.9 83.7 125.5
51 32.1 26.4 39.6
86 64.6 52.8 79.2









#21
trans-1,2-Dichloroethene
Concen: 19.12 ug/l
RT: 5.473 min Scan# 604
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 96 Resp: 98012
Ion Ratio Lower Upper
96 100
61 120.5 98.8 148.2
98 56.0 51.4 77.2



#22
Diisopropyl ether
Concen: 19.37 ug/l
RT: 6.361 min Scan# 755
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 45 Resp: 194273
Ion Ratio Lower Upper
45 100
43 55.0 38.7 58.1
87 33.7 25.6 38.4
59 14.6 9.2 13.8#









#23
 Vinyl Acetate
 Concen: 89.76 ug/l
 RT: 6.303 min Scan# 745
 Delta R.T. -0.006 min
 Lab File: VD069295.D
 Acq: 02 Jun 2021 12:15




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion: 43 Resp: 525080
 Ion Ratio Lower Upper
 43 100
 86 12.9 10.5 15.7



#24
 1,1-Dichloroethane
 Concen: 19.19 ug/l
 RT: 6.261 min Scan# 738
 Delta R.T. -0.006 min
 Lab File: VD069295.D
 Acq: 02 Jun 2021 12:15


Tgt Ion: 63 Resp: 161055
 Ion Ratio Lower Upper
 63 100
 98 8.7 3.8 11.4
 100 5.3 2.6 7.8








#25
2-Butanone
Concen: 94.99 ug/l
RT: 7.208 min Scan# 899
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Instrument :
MSVOA_D
ClientSampled :







Tgt Ion: 43 Resp: 90293
Ion Ratio Lower Upper
43 100
72 31.9 23.9 35.9



#26
2,2-Dichloropropane
Concen: 18.95 ug/l
RT: 7.208 min Scan# 899
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Tgt Ion: 77 Resp: 146541
Ion Ratio Lower Upper
77 100
97 22.5 12.1 36.3








#27
 cis-1,2-Dichloroethene
 Concen: 19.05 ug/l
 RT: 7.208 min Scan# 899
 Delta R.T. -0.006 min
 Lab File: VD069295.D
 Acq: 02 Jun 2021 12:15

Instrument :
 MSVOA_D
 ClientSampleId :



Tgt Ion: 96 Resp: 103385




Ion	Ratio	Lower	Upper
96	100		
61	129.9	0.0	265.6
98	63.4	0.0	129.0




#28
 Bromochloromethane
 Concen: 20.86 ug/l
 RT: 7.544 min Scan# 956
 Delta R.T. -0.000 min
 Lab File: VD069295.D
 Acq: 02 Jun 2021 12:15

Tgt Ion: 49 Resp: 60674

Ion	Ratio	Lower	Upper
49	100		
129	1.0	0.0	4.6
130	89.6	72.0	108.0





#29

Tetrahydrofuran

Concen: 90.65 ug/l

RT: 7.561 min Scan# 959





Delta R.T. -0.006 min

Lab File: VD069295.D

Acq: 02 Jun 2021 12:15

Tgt Ion: 42 Resp: 49528

Ion	Ratio	Lower	Upper
42	100		
72	50.9	41.0	61.6
71	48.8	38.9	58.3



#30

Chloroform

Concen: 19.36 ug/l

RT: 7.708 min Scan# 984




Delta R.T. -0.000 min


Lab File: VD069295.D

Acq: 02 Jun 2021 12:15

Tgt Ion: 83 Resp: 177574


Ion	Ratio	Lower	Upper
83	100		
85	63.8	52.6	78.8








#31
Cyclohexane
Concen: 18.04 ug/l
RT: 7.985 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Instrument :
MSVOA_D
ClientSampled :





Tgt Ion: 56 Resp: 118672
Ion Ratio Lower Upper
56 100
69 43.5 29.0 43.4#
84 106.3 81.1 121.7




#32
1,1,1-Trichloroethane
Concen: 19.00 ug/l
RT: 7.903 min Scan# 1017
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15



Tgt Ion: 97 Resp: 157274
Ion Ratio Lower Upper
97 100
99 66.2 51.9 77.9
61 41.2 31.8 47.6









#33
1,2-Dichloroethane-d4
Concen: 44.45 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 65 Resp: 184000
Ion Ratio Lower Upper
65 100
67 53.6 0.0 105.8



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.867 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 114 Resp: 689261
Ion Ratio Lower Upper
114 100
63 18.6 0.0 35.6
88 14.9 0.0 31.2









#35
Dibromofluoromethane
Concen: 46.09 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:113 Resp: 199503
Ion Ratio Lower Upper
113 100
111 103.6 83.0 124.6
192 18.8 15.1 22.7



#36
1,1-Dichloropropene
Concen: 19.84 ug/l
RT: 8.114 min Scan# 1053
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 75 Resp: 132899
Ion Ratio Lower Upper
75 100
110 36.7 18.4 55.2
77 30.8 24.9 37.3









#37
Ethyl Acetate
Concen: 19.20 ug/l
RT: 7.297 min Scan# 914
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43	Resp: 43678
Ion Ratio	Lower Upper
43 100	
61 15.2	11.0 16.6
70 11.8	9.7 14.5



#38
Carbon Tetrachloride
Concen: 19.84 ug/l
RT: 8.097 min Scan# 1050
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:117	Resp:	143486	
Ion	Ratio	Lower	Upper
117	100		
119	94.8	79.4	119.2
121	30.9	26.1	39.1





#39

Methylcyclohexane

Concen: 19.36 ug/l

RT: 9.355 min Scan# 1264





Delta R.T. -0.000 min

Lab File: VD069295.D

Acq: 02 Jun 2021 12:15

Tgt Ion: 83 Resp: 132684

Ion	Ratio	Lower	Upper
83	100		
55	67.8	57.4	86.2
98	50.9	42.2	63.2



#40

Benzene

Concen: 19.45 ug/l

RT: 8.350 min Scan# 1093




Delta R.T. -0.006 min


Lab File: VD069295.D

Acq: 02 Jun 2021 12:15

Tgt Ion: 78 Resp: 378394

Ion	Ratio	Lower	Upper
78	100		
77	23.9	18.7	28.1









#41
Methacrylonitrile
Concen: 21.02 ug/l
RT: 7.526 min Scan# 953
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion: 41 Resp: 23210




Ion	Ratio	Lower	Upper
41	100		
39	61.9	50.6	76.0
67	86.5	86.2	129.4
52	33.4	32.3	48.5




#42
1,2-Dichloroethane
Concen: 19.69 ug/l
RT: 8.426 min Scan# 1106
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 62 Resp: 102262


Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	21.4








#43
Isopropyl Acetate
Concen: 18.92 ug/l
RT: 8.450 min Scan# 1110
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion: 43 Resp: 78332
Ion Ratio Lower Upper
43 100
61 36.3 29.0 43.6
87 16.1 12.7 19.1




#44
Trichloroethene
Concen: 19.74 ug/l
RT: 9.114 min Scan# 1223
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15



Tgt Ion:130 Resp: 104353
Ion Ratio Lower Upper
130 100
95 94.8 0.0 198.8





#45

1,2-Dichloropropane

Concen: 20.05 ug/l

RT: 9.385 min Scan# 1269

Delta R.T. -0.000 min

Lab File: VD069295.D





Acq: 02 Jun 2021 12:15

Tgt Ion: 63 Resp: 92879

Ion Ratio Lower Upper

63 100

65 30.8 25.3 37.9



#46

Dibromomethane

Concen: 19.41 ug/l

RT: 9.473 min Scan# 1284

Delta R.T. -0.006 min

Lab File: VD069295.D

Acq: 02 Jun 2021 12:15




Tgt Ion: 93 Resp: 51395


Ion Ratio Lower Upper

93 100

95 84.2 63.4 95.2

174 97.8 74.5 111.7









#47
Bromodichloromethane
Concen: 19.81 ug/l
RT: 9.661 min Scan# 1316
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 135274
Ion Ratio Lower Upper
83 100
85 61.0 48.2 72.2
127 7.3 7.1 10.7



#48
Methyl methacrylate
Concen: 17.71 ug/l
RT: 9.461 min Scan# 1282
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 41 Resp: 34045
Ion Ratio Lower Upper
41 100
69 115.6 77.6 116.4
39 65.3 45.3 67.9





#49

1,4-Dioxane

Concen: 362.24 ug/l

RT: 9.461 min Scan# 1282

Delta R.T. -0.000 min

Lab File: VD069295.D

Acq: 02 Jun 2021 12:15





Tgt Ion: 88 Resp: 9422

Ion Ratio Lower Upper

88 100

43 29.5 25.4 38.2

58 75.4 50.2 75.4#



#50

Toluene-d8

Concen: 46.27 ug/l

RT: 10.338 min Scan# 1431

Delta R.T. -0.000 min

Lab File: VD069295.D




Acq: 02 Jun 2021 12:15


Tgt Ion: 98 Resp: 728267

Ion Ratio Lower Upper

98 100

100 65.2 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 97.42 ug/l
RT: 10.226 min Scan# 1412
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 43 Resp: 208326
Ion Ratio Lower Upper
43 100
58 42.3 35.1 52.7



#52
Toluene
Concen: 19.84 ug/l
RT: 10.402 min Scan# 1442
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 92 Resp: 243708
Ion Ratio Lower Upper
92 100
91 173.3 138.4 207.6









#53
t-1,3-Dichloropropene
Concen: 19.36 ug/l
RT: 10.620 min Scan# 1479
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 75 Resp: 118050
Ion Ratio Lower Upper
75 100
77 32.1 25.1 37.7



#54
cis-1,3-Dichloropropene
Concen: 19.46 ug/l
RT: 10.085 min Scan# 1388
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15


Tgt Ion: 75 Resp: 141469
Ion Ratio Lower Upper
75 100
77 30.0 26.2 39.2
39 33.5 29.7 44.5








#55
1,1,2-Trichloroethane
Concen: 20.06 ug/l
RT: 10.796 min Scan# 1509
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:	97	Resp:	73694
Ion	Ratio	Lower	Upper
97	100		
83	84.7	69.4	104.0
85	53.6	44.5	66.7
99	63.3	50.7	76.1



#56
Ethyl methacrylate
Concen: 18.42 ug/l
RT: 10.661 min Scan# 1486
Delta R.T. 0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:	69	Resp:	69798
Ion	Ratio	Lower	Upper
69	100		
41	57.0	45.4	68.2
39	28.3	21.7	32.5









#57
1,3-Dichloropropane
Concen: 20.36 ug/l
RT: 10.944 min Scan# 1534
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 76 Resp: 121265
Ion Ratio Lower Upper
76 100
78 32.2 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 109.54 ug/l
RT: 9.944 min Scan# 1364
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 63 Resp: 188398
Ion Ratio Lower Upper
63 100
106 31.3 24.8 37.2









#59
2-Hexanone
Concen: 98.93 ug/l
RT: 10.985 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 43 Resp: 141406
Ion Ratio Lower Upper
43 100
58 59.7 30.8 92.3



#60
Dibromochloromethane
Concen: 19.65 ug/l
RT: 11.138 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 129 Resp: 90923
Ion Ratio Lower Upper
129 100
127 77.4 39.6 118.7









#61
1,2-Dibromoethane
Concen: 20.07 ug/l
RT: 11.244 min Scan# 1585
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:107 Resp: 70329
Ion Ratio Lower Upper
107 100
109 93.2 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 47.38 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 95 Resp: 244379
Ion Ratio Lower Upper
95 100
174 82.5 0.0 162.8
176 78.4 0.0 154.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.643 min Scan# 1653
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:117 Resp: 652292
Ion Ratio Lower Upper
117 100
82 52.1 43.3 64.9
119 32.1 25.3 37.9



#64
Tetrachloroethene
Concen: 19.33 ug/l
RT: 10.873 min Scan# 1522
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:164 Resp: 87766
Ion Ratio Lower Upper
164 100
166 129.7 100.2 150.4
129 101.0 78.1 117.1
131 92.8 76.7 115.1









#65
Chlorobenzene
Concen: 19.89 ug/l
RT: 11.667 min Scan# 1657
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:112 Resp: 265170
Ion Ratio Lower Upper
112 100
114 30.0 25.6 38.4



#66
1,1,1,2-Tetrachloroethane
Concen: 19.28 ug/l
RT: 11.738 min Scan# 1669
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:131 Resp: 96291
Ion Ratio Lower Upper
131 100
133 96.6 47.4 142.2
119 67.4 32.8 98.3









#67
Ethyl Benzene
Concen: 19.42 ug/l
RT: 11.743 min Scan# 1670
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 444750
Ion Ratio Lower Upper
91 100
106 30.7 25.9 38.9



#68
m/p-Xylenes
Concen: 39.52 ug/l
RT: 11.855 min Scan# 1689
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:106 Resp: 354679
Ion Ratio Lower Upper
106 100
91 198.8 158.3 237.5









#69
o-Xylene
Concen: 19.71 ug/l
RT: 12.179 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:106 Resp: 156436
Ion Ratio Lower Upper
106 100
91 208.4 105.5 316.4



#70
Styrene
Concen: 19.79 ug/l
RT: 12.191 min Scan# 1746
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:104 Resp: 272618
Ion Ratio Lower Upper
104 100
78 47.5 37.5 56.3
103 55.8 44.3 66.5









#71
Bromoform
Concen: 20.27 ug/l
RT: 12.355 min Scan# 1774
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:173 Resp: 52498
Ion Ratio Lower Upper
173 100
175 47.7 23.8 71.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:152 Resp: 322925
Ion Ratio Lower Upper
152 100
115 58.4 29.4 88.3
150 164.1 0.0 355.0









#73
Isopropylbenzene
Concen: 19.00 ug/l
RT: 12.473 min Scan# 1794
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:105 Resp: 419815
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.5



#74
N-ethyl acetate
Concen: 18.08 ug/l
RT: 12.285 min Scan# 1762
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 43 Resp: 70553
Ion Ratio Lower Upper
43 100
70 52.8 40.9 61.3
55 30.5 23.4 35.0
61 26.6 23.2 34.8









#75
1,1,2,2-Tetrachloroethane
Concen: 19.31 ug/l
RT: 12.726 min Scan# 1837
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 79657
Ion Ratio Lower Upper
83 100
131 10.9 5.3 16.1
85 63.7 32.6 98.0



#76
1,2,3-Trichloropropane
Concen: 35.37 ug/l
RT: 12.779 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 75 Resp: 99237
Ion Ratio Lower Upper
75 100
77 0.0 183.8 551.2#









#77
Bromobenzene
Concen: 19.60 ug/l
RT: 12.761 min Scan# 1843
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:156 Resp: 104308
Ion Ratio Lower Upper
156 100
77 178.5 89.3 267.9
158 97.1 48.0 144.2



#78
n-propylbenzene
Concen: 19.45 ug/l
RT: 12.814 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 91 Resp: 527213
Ion Ratio Lower Upper
91 100
120 22.9 11.6 34.8









#79
2-Chlorotoluene
Concen: 19.41 ug/l
RT: 12.902 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 298998
Ion Ratio Lower Upper
91 100
126 34.9 16.8 50.3



#80
1,3,5-Trimethylbenzene
Concen: 19.50 ug/l
RT: 12.955 min Scan# 1876
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 105 Resp: 365037
Ion Ratio Lower Upper
105 100
120 50.2 24.9 74.7









#81
trans-1,4-Dichloro-2-butene
Concen: 19.29 ug/l
RT: 12.526 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 24426
Ion Ratio Lower Upper
75 100
53 78.3 62.6 94.0
89 43.7 35.6 53.4



#82
4-Chlorotoluene
Concen: 19.71 ug/l
RT: 13.002 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 91 Resp: 319455
Ion Ratio Lower Upper
91 100
126 34.7 17.4 52.2









#83
tert-Butylbenzene
Concen: 19.11 ug/l
RT: 13.220 min Scan# 1921
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:119 Resp: 299590
Ion Ratio Lower Upper
119 100
91 62.4 31.7 95.1
134 24.7 13.4 40.1



#84
1,2,4-Trimethylbenzene
Concen: 19.84 ug/l
RT: 13.267 min Scan# 1929
Delta R.T. 0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:105 Resp: 369353
Ion Ratio Lower Upper
105 100
120 45.7 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 19.53 ug/l
 RT: 13.396 min Scan# 1951
 Delta R.T. -0.000 min
 Lab File: VD069295.D
 Acq: 02 Jun 2021 12:15




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion:105 Resp: 451920
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.1 30.3



#86
 p-Isopropyltoluene
 Concen: 19.32 ug/l
 RT: 13.508 min Scan# 1970
 Delta R.T. -0.000 min
 Lab File: VD069295.D
 Acq: 02 Jun 2021 12:15

Tgt Ion:119 Resp: 390867
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.0 39.0
 91 22.9 11.3 33.8









#87
1,3-Dichlorobenzene
Concen: 19.74 ug/l
RT: 13.508 min Scan# 1970
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:146 Resp: 212596
Ion Ratio Lower Upper
146 100
111 39.5 20.3 60.9
148 63.3 31.9 95.9



#88
1,4-Dichlorobenzene
Concen: 19.45 ug/l
RT: 13.590 min Scan# 1984
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:146 Resp: 208768
Ion Ratio Lower Upper
146 100
111 40.6 19.4 58.4
148 64.4 31.1 93.3









#89
n-Butylbenzene
Concen: 18.94 ug/l
RT: 13.837 min Scan# 2026
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 358488
Ion Ratio Lower Upper
91 100
92 52.6 26.4 79.2
134 25.2 13.0 38.9



#90
Hexachloroethane
Concen: 19.17 ug/l
RT: 14.102 min Scan# 2071
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 117 Resp: 73684
Ion Ratio Lower Upper
117 100
201 73.3 36.0 108.1









#91
1,2-Dichlorobenzene
Concen: 19.27 ug/l
RT: 13.885 min Scan# 2034
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:146 Resp: 178833
Ion Ratio Lower Upper
146 100
111 41.4 20.5 61.6
148 64.2 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.62 ug/l
RT: 14.496 min Scan# 2138
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion: 75 Resp: 12240
Ion Ratio Lower Upper
75 100
155 80.2 42.1 126.4
157 112.3 54.4 163.3









#93
1,2,4-Trichlorobenzene
Concen: 19.15 ug/l
RT: 15.155 min Scan# 2250
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:180 Resp: 106641
Ion Ratio Lower Upper
180 100
182 94.6 47.3 141.8
145 31.3 15.2 45.5



#94
Hexachlorobutadiene
Concen: 19.02 ug/l
RT: 15.261 min Scan# 2268
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:225 Resp: 66491
Ion Ratio Lower Upper
225 100
223 61.5 30.5 91.5
227 64.1 31.4 94.2









#95
Naphthalene
Concen: 17.88 ug/l
RT: 15.390 min Scan# 2290
Delta R.T. -0.006 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion:128 Resp: 171973
Ion Ratio Lower Upper
128 100
127 13.8 10.4 15.6
129 11.6 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 18.61 ug/l
RT: 15.584 min Scan# 2323
Delta R.T. -0.000 min
Lab File: VD069295.D
Acq: 02 Jun 2021 12:15

Tgt Ion:180 Resp: 91189
Ion Ratio Lower Upper
180 100
182 95.5 47.0 141.0
145 32.4 15.9 47.7

