

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069300.D
 Acq On : 02 Jun 2021 14:36
 Operator : VA/SY
 Sample : M2580-01
 Misc : 5.67G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 18-B1(0-2)

Quant Time: Jun 03 01:08:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	266680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	555809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	544872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	249097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	187715	73.359	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 146.720%	
35) Dibromofluoromethane	7.914	113	201422	57.708	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.420%	
50) Toluene-d8	10.338	98	706760	55.685	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.360%	
62) 4-Bromofluorobenzene	12.632	95	255058	61.328	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 122.660%	
Target Compounds						
95) Naphthalene	15.390	128	7802	4.159	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
Data File : VD069300.D
Acq On : 02 Jun 2021 14:36
Operator : VA/SY
Sample : M2580-01
Misc : 5.67G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
18-B1(0-2)

Quant Time: Jun 03 01:08:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

