

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069305.D
 Acq On : 02 Jun 2021 16:58
 Operator : VA/SY
 Sample : M2580-06
 Misc : 15.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 18-B2(4-6)

Quant Time: Jun 03 01:10:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	297613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	616124	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	623149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	249551	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	217314	76.100	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 152.200%			
35) Dibromofluoromethane	7.920	113	231385	59.803	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 119.600%			
50) Toluene-d8	10.337	98	806177	57.300	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 114.600%			
62) 4-Bromofluorobenzene	12.631	95	268565	58.255	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 116.500%			
Target Compounds					Qvalue	
16) Acetone	4.161	43	76127	152.763	ug/l	99
17) Carbon Disulfide	4.408	76	34919	3.538	ug/l #	95
25) 2-Butanone	7.220	43	16570	25.269	ug/l	98
95) Naphthalene	15.390	128	13384	4.770	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
Data File : VD069305.D
Acq On : 02 Jun 2021 16:58
Operator : VA/SY
Sample : M2580-06
Misc : 15.62G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
18-B2(4-6)

Quant Time: Jun 03 01:10:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

