

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069310.D
 Acq On : 02 Jun 2021 19:19
 Operator : VA/SY
 Sample : M2580-10
 Misc : 8.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 18-B4(0-2)

Quant Time: Jun 03 01:11:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	282996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	600849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	592671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	246624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	202217	74.470	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	148.940%	
35) Dibromofluoromethane	7.914	113	216417	57.357	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	114.720%	
50) Toluene-d8	10.338	98	770627	56.165	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	112.340%	
62) 4-Bromofluorobenzene	12.632	95	250834	55.792	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	111.580%	
Target Compounds						
95) Naphthalene	15.391	128	7964	4.186	ug/l	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
Data File : VD069310.D
Acq On : 02 Jun 2021 19:19
Operator : VA/SY
Sample : M2580-10
Misc : 8.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
18-B4(0-2)

Quant Time: Jun 03 01:11:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

