

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069293.D
 Acq On : 02 Jun 2021 11:09
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 03 01:06:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.526	0.485	7.8	90	0.00
3 P	Chloromethane	0.397	0.375	5.5	98	0.00
4 C	Vinyl Chloride	0.405	0.401	1.0#	99	0.00
5 T	Bromomethane	0.295	0.303	-2.7	106	0.00
6 T	Chloroethane	0.231	0.230	0.4	102	0.00
7 T	Trichlorofluoromethane	0.942	0.892	5.3	93	0.00
8 T	Diethyl Ether	0.231	0.224	3.0	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.554	0.532	4.0	97	0.00
10 T	Methyl Iodide	0.503	0.479	4.8	81	0.00
11 T	Tert butyl alcohol	0.038	0.025	34.2#	94	0.00
12 CM	1,1-Dichloroethene	0.501	0.477	4.8#	92	0.00
13 T	Acrolein	0.028	0.029	-3.6	86	0.00
14 T	Allyl chloride	0.605	0.574	5.1	91	0.00
15 T	Acrylonitrile	0.090	0.085	5.6	91	0.00
16 T	Acetone	0.084	0.084	0.0	105	0.00
17 T	Carbon Disulfide	1.658	1.591	4.0	92	0.00
18 T	Methyl Acetate	0.227	0.191	15.9	87	0.00
19 T	Methyl tert-butyl Ether	0.921	0.903	2.0	93	0.00
20 T	Methylene Chloride	0.778	0.613	21.2	99	0.00
21 T	trans-1,2-Dichloroethene	0.594	0.580	2.4	95	0.00
22 T	Diisopropyl ether	1.162	1.167	-0.4	93	0.00
23 T	Vinyl Acetate	0.669	0.694	-3.7	92	0.00
24 P	1,1-Dichloroethane	0.972	0.931	4.2	96	0.00
25 T	2-Butanone	0.110	0.106	3.6	95	0.00
26 T	2,2-Dichloropropane	0.896	0.831	7.3	95	0.00
27 T	cis-1,2-Dichloroethene	0.629	0.621	1.3	96	0.00
28 T	Bromochloromethane	0.337	0.324	3.9	101	0.00
29 T	Tetrahydrofuran	0.063	0.060	4.8	90	0.00
30 C	Chloroform	1.063	1.015	4.5#	96	0.00
31 T	Cyclohexane	0.762	0.690	9.4	90	0.00
32 T	1,1,1-Trichloroethane	0.959	0.917	4.4	96	0.00
33 S	1,2-Dichloroethane-d4	0.480	0.408	15.0	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.314	0.283	9.9	99	0.00
36 T	1,1-Dichloropropene	0.486	0.487	-0.2	94	0.00
37 T	Ethyl Acetate	0.165	0.156	5.5	90	0.00
38 T	Carbon Tetrachloride	0.525	0.520	1.0	95	0.00
39 T	Methylcyclohexane	0.497	0.518	-4.2	92	0.00
40 TM	Benzene	1.412	1.415	-0.2	96	0.00
41 T	Methacrylonitrile	0.080	0.084	-5.0	106	0.00
42 TM	1,2-Dichloroethane	0.377	0.368	2.4	95	0.00
43 T	Isopropyl Acetate	0.300	0.288	4.0	91	0.00
44 TM	Trichloroethene	0.384	0.385	-0.3	96	0.00
45 C	1,2-Dichloropropane	0.336	0.344	-2.4#	98	0.00
46 T	Dibromomethane	0.192	0.191	0.5	93	0.00
47 T	Bromodichloromethane	0.495	0.493	0.4	97	0.00
48 T	Methyl methacrylate	0.139	0.135	2.9	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
 Data File : VD069293.D
 Acq On : 02 Jun 2021 11:09
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 03 01:06:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	98	0.00
50 S	Toluene-d8	1.142	1.013	11.3	96	0.00
51 T	4-Methyl-2-Pentanone	0.155	0.154	0.6	91	0.00
52 CM	Toluene	0.891	0.928	-4.2#	97	0.00
53 T	t-1,3-Dichloropropene	0.442	0.441	0.2	94	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.521	1.1	95	0.00
55 T	1,1,2-Trichloroethane	0.266	0.261	1.9	95	0.00
56 T	Ethyl methacrylate	0.274	0.284	-3.6	93	0.00
57 T	1,3-Dichloropropane	0.432	0.432	0.0	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.139	-11.2	114	0.00
59 T	2-Hexanone	0.104	0.108	-3.8	95	0.00
60 T	Dibromochloromethane	0.336	0.339	-0.9	99	0.00
61 T	1,2-Dibromoethane	0.254	0.254	0.0	97	0.00
62 S	4-Bromofluorobenzene	0.374	0.341	8.8	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.348	0.338	2.9	96	0.00
65 PM	Chlorobenzene	1.022	1.017	0.5	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.383	0.378	1.3	97	0.00
67 C	Ethyl Benzene	1.755	1.823	-3.9#	96	0.00
68 T	m/p-Xylenes	0.688	0.719	-4.5	95	0.00
69 T	o-Xylene	0.608	0.648	-6.6	96	0.00
70 T	Styrene	1.056	1.140	-8.0	96	0.00
71 P	Bromoform	0.199	0.202	-1.5	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	3.422	3.469	-1.4	95	0.00
74 T	N-amyl acetate	0.604	0.575	4.8	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.639	0.593	7.2	93	0.00
76 T	1,2,3-Trichloropropane	0.434	0.386	11.1	95	0.00
77 T	Bromobenzene	0.824	0.805	2.3	96	0.00
78 T	n-propylbenzene	4.197	4.282	-2.0	95	0.00
79 T	2-Chlorotoluene	2.386	2.349	1.6	94	0.00
80 T	1,3,5-Trimethylbenzene	2.898	2.983	-2.9	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.196	0.184	6.1	89	0.00
82 T	4-Chlorotoluene	2.510	2.488	0.9	96	0.00
83 T	tert-Butylbenzene	2.427	2.441	-0.6	93	0.00
84 T	1,2,4-Trimethylbenzene	2.882	2.979	-3.4	96	0.00
85 T	sec-Butylbenzene	3.582	3.639	-1.6	93	0.00
86 T	p-Isopropyltoluene	3.132	3.251	-3.8	95	0.00
87 T	1,3-Dichlorobenzene	1.668	1.633	2.1	97	0.00
88 T	1,4-Dichlorobenzene	1.662	1.607	3.3	97	0.00
89 T	n-Butylbenzene	2.931	2.960	-1.0	92	0.00
90 T	Hexachloroethane	0.595	0.561	5.7	93	0.00
91 T	1,2-Dichlorobenzene	1.437	1.409	1.9	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.090	11.8	93	0.00
93 T	1,2,4-Trichlorobenzene	0.862	0.862	0.0	96	0.00
94 T	Hexachlorobutadiene	0.541	0.514	5.0	93	0.00
95 T	Naphthalene	1.520	1.477	2.8	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060221\
Data File : VD069293.D
Acq On : 02 Jun 2021 11:09
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 03 01:06:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.759	0.753	0.8	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6