

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060223\
 Data File : VD076318.D
 Acq On : 02 Jun 2023 17:33
 Operator : KP/SY
 Sample : 02986-04
 Misc : 5.61G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-14-0A

Quant Time: Jun 03 02:24:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 05:50:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	166826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	314658	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	288578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	118147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	99361	61.115	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.220%
35) Dibromofluoromethane	7.799	113	123347	59.384	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.760%
50) Toluene-d8	10.269	98	385721	48.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.320%
62) 4-Bromofluorobenzene	12.575	95	102545	45.908	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.820%
Target Compounds						
16) Acetone	4.017	43	11227	27.278	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060223\
Data File : VD076318.D
Acq On : 02 Jun 2023 17:33
Operator : KP/SY
Sample : 02986-04
Misc : 5.61G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-14-0A

Quant Time: Jun 03 02:24:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 05:50:45 2023
Response via : Initial Calibration

