

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060421\
 Data File : VD069352.D
 Acq On : 04 Jun 2021 18:12
 Operator : VA/SY
 Sample : M2598-07
 Misc : 6.20G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 13

Quant Time: Jun 05 00:37:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	202133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	357898	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	374272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	173351	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	147066	75.827	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.660%
35) Dibromofluoromethane	7.914	113	137583	61.216	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.440%
50) Toluene-d8	10.338	98	443434	54.257	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.520%
62) 4-Bromofluorobenzene	12.632	95	149443	55.804	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060421\
Data File : VD069352.D
Acq On : 04 Jun 2021 18:12
Operator : VA/SY
Sample : M2598-07
Misc : 6.20G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
13

Quant Time: Jun 05 00:37:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

