

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060519\
 Data File : VD062617.D
 Acq On : 5 Jun 2019 17:41
 Operator : FY/SY
 Sample : VSTDICV050
 Misc : 5.00µ/5mL/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519

Quant Time: Jun 06 05:54:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 06 05:42:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	825766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.24	114	1117862	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.38	117	937780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.53	152	517115	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.47	65	559588	49.43	ug/l	0.00
Spiked Amount			Recovery	=	98.86%	
35) Dibromofluoromethane	6.94	113	567650	49.96	ug/l	0.00
Spiked Amount			Recovery	=	99.92%	
50) Toluene-d8	10.42	98	1264485	51.66	ug/l	0.00
Spiked Amount			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	13.56	95	570729	49.46	ug/l	0.00
Spiked Amount			Recovery	=	98.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.58	85	410432	44.499	ug/l	97
3) Chloromethane	1.75	50	347269	43.545	ug/l	99
4) Vinyl Chloride	1.86	62	385243	49.266	ug/l	96
5) Bromomethane	2.16	94	204689	51.887	ug/l	96
6) Chloroethane	2.28	64	200834	50.920	ug/l	95
7) Trichlorofluoromethane	2.54	101	858727	49.222	ug/l	98
8) Diethyl Ether	2.88	74	124256	54.805	ug/l	92
9) 1,1,2-Trichlorotrifluoroet	3.16	101	486534	49.604	ug/l	96
10) Methyl Iodide	3.31	142	646951	51.532	ug/l	95
11) Tert butyl alcohol	4.09	59	124047	244.728	ug/l #	90
12) 1,1-Dichloroethene	3.14	96	347688	48.317	ug/l	83
13) Acrolein	3.05	56	41538	231.742	ug/l	98
14) Allyl chloride	3.63	41	572790	52.284	ug/l	96
15) Acrylonitrile	4.21	53	274958	248.838	ug/l #	83
16) Acetone	3.24	43	571775	281.067	ug/l #	90
17) Carbon Disulfide	3.38	76	1105763	52.739	ug/l	95
18) Methyl Acetate	3.65	43	167902	46.420	ug/l	97
19) Methyl tert-butyl Ether	4.26	73	861435	49.422	ug/l	97
20) Methylene Chloride	3.83	84	359460	45.402	ug/l	95
21) trans-1,2-Dichloroethene	4.23	96	395033	50.446	ug/l	97
22) Diisopropyl ether	5.14	45	1161536	50.190	ug/l	99
23) Vinyl Acetate	5.08	43	3768296	267.162	ug/l	99
24) 1,1-Dichloroethane	5.00	63	741643	50.863	ug/l	99
25) 2-Butanone	6.09	43	523299	270.801	ug/l	96
26) 2,2-Dichloropropane	6.04	77	839289	52.686	ug/l	94
27) cis-1,2-Dichloroethene	6.05	96	430353	50.965	ug/l	92
28) Bromochloromethane	6.45	49	317995	57.642	ug/l	95
29) Tetrahydrofuran	6.48	42	225188	251.033	ug/l	96
30) Chloroform	6.68	83	922593	50.873	ug/l	95
31) Cyclohexane	6.97	56	504579	54.132	ug/l	96
32) 1,1,1-Trichloroethane	6.89	97	935723	50.974	ug/l	98
36) 1,1-Dichloropropene	7.16	75	597740	49.734	ug/l	96
37) Ethyl Acetate	6.20	43	245511	50.620	ug/l	96
38) Carbon Tetrachloride	7.12	117	992228	55.005	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060519\
 Data File : VD062617.D
 Acq On : 5 Jun 2019 17:41
 Operator : FY/SY
 Sample : VSTDICV050
 Misc : 5.00µ/5mL/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICSVVD060519

Quant Time: Jun 06 05:54:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 06 05:42:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	8.88	83	512997	49.923	ug/l	98
40) Benzene	7.47	78	1162449	48.970	ug/l	100
41) Methacrylonitrile	6.46	41	325275	51.788	ug/l	98
42) 1,2-Dichloroethane	7.60	62	702740	50.295	ug/l	100
43) Isopropyl Acetate	9.25	43	352068	52.814	ug/l	100
44) Trichloroethene	8.55	130	478933	49.753	ug/l	98
45) 1,2-Dichloropropane	8.96	63	301096	50.655	ug/l	91
46) Dibromomethane	9.07	93	259317	50.146	ug/l	99
47) Bromodichloromethane	9.39	83	722690	51.885	ug/l	97
48) Methyl methacrylate	9.13	41	223526	49.305	ug/l	93
49) 1,4-Dioxane	9.09	88	36064	1002.342	ug/l #	92
51) 4-Methyl-2-Pentanone	10.31	43	1173372	270.323	ug/l	98
52) Toluene	10.52	92	826947	52.522	ug/l	97
53) t-1,3-Dichloropropene	10.92	75	627637	54.270	ug/l	99
54) cis-1,3-Dichloropropene	10.05	75	644598	54.733	ug/l	98
55) 1,1,2-Trichloroethane	11.20	97	259649	48.068	ug/l	99
56) Ethyl methacrylate	11.05	69	305702	50.780	ug/l	96
57) 1,3-Dichloropropane	11.42	76	442701	52.493	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.87	63	682965	221.190	ug/l	97
59) 2-Hexanone	11.54	43	852024	258.874	ug/l	95
60) Dibromochloromethane	11.69	129	546732	51.799	ug/l	96
61) 1,2-Dibromoethane	11.82	107	346127	52.318	ug/l	100
64) Tetrachloroethene	11.27	164	427871	50.520	ug/l	97
65) Chlorobenzene	12.41	112	1000627	51.350	ug/l	99
66) 1,1,1,2-Tetrachloroethane	12.52	131	473336	50.532	ug/l	98
67) Ethyl Benzene	12.53	91	1783157	50.629	ug/l	98
68) m/p-Xylenes	12.68	106	1249263	101.688	ug/l	95
69) o-Xylene	13.06	106	578447	48.868	ug/l	100
70) Styrene	13.08	104	1054738	53.040	ug/l	99
71) Bromoform	13.25	173	330022	51.011	ug/l #	95
73) Isopropylbenzene	13.41	105	1833652	50.996	ug/l	97
74) N-amyl acetate	13.28	43	472391	50.880	ug/l	92
75) 1,1,2,2-Tetrachloroethane	13.71	83	310183	51.461	ug/l	96
76) 1,2,3-Trichloropropane	13.74	75	337056	49.967	ug/l	98
77) Bromobenzene	13.68	156	517920	51.785	ug/l	84
78) n-propylbenzene	13.78	91	2014777	48.316	ug/l	99
79) 2-Chlorotoluene	13.85	91	1279540	50.612	ug/l	98
80) 1,3,5-Trimethylbenzene	13.94	105	1605419	51.880	ug/l	98
81) trans-1,4-Dichloro-2-buten	13.48	75	98901	57.039	ug/l	87
82) 4-Chlorotoluene	13.96	91	1564768	52.598	ug/l	97
83) tert-Butylbenzene	14.20	119	1775715	52.113	ug/l	88
84) 1,2,4-Trimethylbenzene	14.24	105	1585306	50.051	ug/l	100
85) sec-Butylbenzene	14.38	105	1916097	54.575	ug/l	95
86) p-Isopropyltoluene	14.50	119	1689286	50.046	ug/l	99
87) 1,3-Dichlorobenzene	14.47	146	864064	51.476	ug/l	95
88) 1,4-Dichlorobenzene	14.55	146	901564	53.033	ug/l	97
89) n-Butylbenzene	14.80	91	1544624	49.214	ug/l	98
90) Hexachloroethane	15.02	117	463865	52.214	ug/l	73
91) 1,2-Dichlorobenzene	14.81	146	730653	47.141	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	15.38	75	62813	50.944	ug/l	83


Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060519\
Data File : VD062617.D
Acq On : 5 Jun 2019 17:41
Operator : FY/SY
Sample : VSTDICV050
Misc : 5.00g/5mL/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Quant Time: Jun 06 05:54:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060519S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 06 05:42:40 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	15.95	180	567042	50.917	ug/l	96
94) Hexachlorobutadiene	16.04	225	444762	53.243	ug/l	98
95) Naphthalene	16.13	128	858976	53.863	ug/l	99
96) 1,2,3-Trichlorobenzene	16.28	180	422881	49.658	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.07 min Scan# 573

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :


ICVVD060519

Tot Ion:168 Resp: 825766

Ion Ratio Lower Upper

168 100

99 58.8 48.1 72.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC

7.07

300000

250000

200000

150000



100000

50000

0

6.90 7.00 7.10 7.20 7.30

Time-->



#2

Dichlorodifluoromethane

Concen: 44.499 ug/l

RT: 1.58 min Scan# 16

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 85 Resp: 410432

Ion Ratio Lower Upper

85 100

87 31.8 16.9 50.6



Abundance Ion 84.95 (84.65 to 85.65): VDC

Ion 86.90 (86.60 to 87.60): VDC

1.58

200000

150000


100000


50000

0

1.50 1.60 1.70 1.80

Time-->





#3

Chloromethane

Concen: 43.545 ug/l

RT: 1.75 min Scan# 33

Delta R.T. -0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :


ICVVD060519

Tot Ion: 50 Resp: 347269

Ion Ratio Lower Upper

50 100

52 32.7 26.6 40.0



Abundance Ion 49.90 (49.60 to 50.60): VDC

Ion 51.90 (51.60 to 52.60): VDC

1.75



100000

50000

0

1.70 1.80 1.90

Time-->



#4

Vinyl Chloride

Concen: 49.266 ug/l

RT: 1.86 min Scan# 44

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 62 Resp: 385243

Ion Ratio Lower Upper

62 100

64 31.9 27.2 40.8



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 63.95 (63.65 to 64.65): VDC

1.86

200000

150000


100000


50000

0

1.80 1.90 2.00

Time-->





#5

Bromomethane

Concen: 51.887 ug/l

RT: 2.16 min Scan# 75

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 94 Resp: 204689

Ion Ratio Lower Upper




94 100

96 99.1 75.8 113.8

93 22.7 16.6 25.0



Abundance Ion 94.00 (93.70 to 94.70): VDC
Ion 96.00 (95.70 to 96.70): VDC
Ion 93.00 (92.70 to 93.70): VDC



#6

Chloroethane

Concen: 50.920 ug/l

RT: 2.28 min Scan# 87

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 64 Resp: 200834



Ion Ratio Lower Upper


64 100

66 32.5 28.4 42.6



Abundance Ion 63.95 (63.65 to 64.65): VDC
Ion 65.90 (65.60 to 66.60): VDC






#7




Trichlorofluoromethane
Concen: 49.222 ug/l
RT: 2.54 min Scan# 113
Delta R.T. -0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion:101 Resp: 858727
Ion Ratio Lower Upper
101 100
103 63.9 52.1 78.1




Abundance Ion 100.85 (100.55 to 101.55): V
Ion 102.85 (102.55 to 103.55): V





#8


Diethyl Ether
Concen: 54.805 ug/l
RT: 2.88 min Scan# 148
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 74 Resp: 124256
Ion Ratio Lower Upper
74 100
45 140.4 75.6 226.8



Abundance Ion 74.00 (73.70 to 74.70): VDC
Ion 45.05 (44.75 to 45.75): VDC





#9

1,1,2-Trichlorotrifluoroethane

Concen: 49.604 ug/l

RT: 3.16 min Scan# 176

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion:101 Resp: 486534

Ion Ratio Lower Upper

101 100

85 44.6 33.8 50.6




151 74.3 56.8 85.2



Abundance Ion 100.85 (100.55 to 101.55): V

Ion 84.95 (84.65 to 85.65): V

Ion 150.90 (150.60 to 151.60): V



#10

Methyl Iodide

Concen: 51.532 ug/l

RT: 3.31 min Scan# 192

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion:142 Resp: 646951

Ion Ratio Lower Upper

142 100

127 67.1 50.7 76.1



141 14.0 12.9 19.3




Abundance Ion 141.85 (141.55 to 142.55): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.85 (140.55 to 141.55): V









#11
Tert butyl alcohol
Concen: 244.728 ug/l
RT: 4.09 min Scan# 271
Delta R.T. -0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 59 Resp: 124047
Ion Ratio Lower Upper
59 100
57 11.6 6.4 9.6#



#12
1,1-Dichloroethene
Concen: 48.317 ug/l
RT: 3.14 min Scan# 174
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 96 Resp: 347688
Ion Ratio Lower Upper
96 100
61 220.0 151.3 226.9
98 63.6 48.6 72.8









#13
Acrolein
Concen: 231.742 ug/l
RT: 3.05 min Scan# 165
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 56 Resp: 41538
Ion Ratio Lower Upper
56 100
55 74.6 61.4 92.0



#14
Allyl chloride
Concen: 52.284 ug/l
RT: 3.63 min Scan# 224
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 41 Resp: 572790
Ion Ratio Lower Upper
41 100
39 114.3 95.5 143.3
76 33.8 25.4 38.0









#15
 Acrylonitrile
 Concen: 248.838 ug/l
 RT: 4.21 min Scan# 283
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleID :
 ICVVD060519


Tot Ion: 53 Resp: 274958
 Ion Ratio Lower Upper
 53 100
 52 102.4 65.8 98.6#
 51 50.0 36.9 55.3



#16
 Acetone
 Concen: 281.067 ug/l
 RT: 3.24 min Scan# 184
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 43 Resp: 571775
 Ion Ratio Lower Upper
 43 100
 58 17.6 10.7 16.1#





#17

Carbon Disulfide

Concen: 52.739 ug/l

RT: 3.38 min Scan# 199

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :


ICVVD060519

Tot Ion: 76 Resp: 1105763

Ion Ratio Lower Upper




76 100

78 8.2 8.1 12.1



Abundance Ion 75.85 (75.55 to 76.55): VDC

Ion 77.85 (77.55 to 78.55): VDC



#18

Methyl Acetate

Concen: 46.420 ug/l

RT: 3.65 min Scan# 226

Delta R.T. -0.01 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 43 Resp: 167902

Ion Ratio Lower Upper



43 100


74 20.9 15.5 23.3



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 74.00 (73.70 to 74.70): VDC









#19

Methyl tert-butyl Ether
 Concen: 49.422 ug/l
 RT: 4.26 min Scan# 288
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519




Tot Ion: 73 Resp: 861435
 Ion Ratio Lower Upper
 73 100
 57 18.1 13.5 20.3




#20

Methylene Chloride
 Concen: 45.402 ug/l
 RT: 3.83 min Scan# 244
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 84 Resp: 359460
 Ion Ratio Lower Upper
 84 100
 49 144.9 119.0 178.6
 51 44.2 35.4 53.2
 86 72.1 50.0 75.0





#21

trans-1,2-Dichloroethene

Concen: 50.446 ug/l

RT: 4.23 min Scan# 285

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 96 Resp: 395033

Ion Ratio Lower Upper




96 100

61 158.1 128.2 192.4

98 54.5 47.5 71.3



Abundance Ion 95.90 (95.60 to 96.60): VDC
Ion 60.95 (60.65 to 61.65): VDC
Ion 97.85 (97.55 to 98.55): VDC



#22

Diisopropyl ether

Concen: 50.190 ug/l

RT: 5.14 min Scan# 377

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Tgt Ion: 45 Resp: 1161536


Ion Ratio Lower Upper

45 100



43 72.8 59.3 88.9


87 18.6 14.6 22.0

59 9.0 6.6 10.0



Abundance Ion 45.05 (44.75 to 45.75): VDC
Ion 43.05 (42.75 to 43.75): VDC
Ion 87.00 (86.70 to 87.70): VDC
Ion 58.95 (58.65 to 59.65): VDC





#23

Vinyl Acetate

Concen: 267.162 ug/l

RT: 5.08 min Scan# 371

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :


ICVVD060519

Tot Ion: 43 Resp: 3768296

Ion Ratio Lower Upper




43 100

86 6.1 5.2 7.8



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 86.00 (85.70 to 86.70): VDC



#24

1,1-Dichloroethane

Concen: 50.863 ug/l

RT: 5.00 min Scan# 363

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion: 63 Resp: 741643

Ion Ratio Lower Upper

63 100

98 4.0 1.7 5.1



100 2.3 1.1 3.5




Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 97.95 (97.65 to 98.65): VDC

Ion 99.95 (99.65 to 100.65): VDC





#25

2-Butanone

Concen: 270.801 ug/l

RT: 6.09 min Scan# 474

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :





ICVVD060519

Tot Ion: 43 Resp: 523299

Ion Ratio Lower Upper

43 100

72 17.9 15.7 23.5



#26

2,2-Dichloropropane

Concen: 52.686 ug/l

RT: 6.04 min Scan# 469

Delta R.T. -0.00 min

Lab File: VD062617.D




Acq: 5 Jun 2019 17:41


Tgt Ion: 77 Resp: 839289

Ion Ratio Lower Upper

77 100

97 16.6 9.6 28.8









#27

cis-1,2-Dichloroethene
Concen: 50.965 ug/l
RT: 6.05 min Scan# 470
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519




Tot Ion: 96 Resp: 430353
Ion Ratio Lower Upper
96 100
61 173.8 0.0 372.0
98 70.0 0.0 126.4




#28

Bromochloromethane
Concen: 57.642 ug/l
RT: 6.45 min Scan# 511
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 49 Resp: 317995
Ion Ratio Lower Upper
49 100
129 2.7 0.0 6.4
130 91.9 77.6 116.4





#29

Tetrahydrofuran

Concen: 251.033 ug/l

RT: 6.48 min Scan# 514

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 42 Resp: 225188

Ion Ratio Lower Upper




42 100

72 45.3 32.5 48.7

71 43.3 35.0 52.4



Abundance Ion 42.00 (41.70 to 42.70): VDC
Ion 72.00 (71.70 to 72.70): VDC
Ion 71.00 (70.70 to 71.70): VDC



#30

Chloroform

Concen: 50.873 ug/l

RT: 6.68 min Scan# 534

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 83 Resp: 922593



Ion Ratio Lower Upper


83 100

85 66.6 50.4 75.6



Abundance Ion 82.95 (82.65 to 83.65): VDC
Ion 84.95 (84.65 to 85.65): VDC









#31
Cyclohexane
Concen: 54.132 ug/l
RT: 6.97 min Scan# 563
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 56 Resp: 504579
Ion Ratio Lower Upper
56 100
69 33.1 23.0 34.4
84 85.1 66.0 99.0



#32
1,1,1-Trichloroethane
Concen: 50.974 ug/l
RT: 6.89 min Scan# 555
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 97 Resp: 935723
Ion Ratio Lower Upper
97 100
99 68.3 53.4 80.0
61 44.2 36.7 55.1









#33
 1,2-Dichloroethane-d4
 Concen: 49.435 ug/l
 RT: 7.47 min Scan# 614
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion: 65 Resp: 559588
 Ion Ratio Lower Upper
 65 100
 67 44.4 0.0 86.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.24 min Scan# 692
 Delta R.T. 0.01 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 114 Resp: 1117862
 Ion Ratio Lower Upper
 114 100
 63 20.1 0.0 41.0
 88 17.9 0.0 34.2





#35

Dibromofluoromethane

Concen: 49.963 ug/l

RT: 6.94 min Scan# 560

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion:113 Resp: 567650

Ion Ratio Lower Upper




113 100

111 109.0 80.1 120.1

192 19.0 13.0 19.6



Abundance Ion 112.90 (112.60 to 113.60): V
Ion 110.90 (110.60 to 111.60): V
Ion 191.80 (191.50 to 192.50): V



#36

1,1-Dichloropropene

Concen: 49.734 ug/l

RT: 7.16 min Scan# 583

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion: 75 Resp: 597740

Ion Ratio Lower Upper



75 100


110 28.8 16.5 49.5

77 28.6 23.3 34.9



Abundance Ion 75.00 (74.70 to 75.70): V
Ion 109.90 (109.60 to 110.60): V
Ion 76.95 (76.65 to 77.65): V









#37
Ethyl Acetate
Concen: 50.620 ug/l
RT: 6.20 min Scan# 485
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 43 Resp: 245511
Ion Ratio Lower Upper
43 100
61 9.4 8.8 13.2
70 8.4 5.6 8.4



#38
Carbon Tetrachloride
Concen: 55.005 ug/l
RT: 7.12 min Scan# 579
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 117 Resp: 992228
Ion Ratio Lower Upper
117 100
119 92.9 74.1 111.1
121 30.2 23.5 35.3





#39

Methvlcvclohexane

Concen: 49.923 ug/l

RT: 8.88 min Scan# 757

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 83 Resp: 512997

Ion Ratio Lower Upper

83 100

55 100.7 83.2 124.8




98 50.2 39.8 59.8



Abundance Ion 83.05 (82.75 to 83.75): VDC

Ion 55.00 (54.70 to 55.70): VDC

Ion 98.05 (97.75 to 98.75): VDC



#40

Benzene

Concen: 48.970 ug/l

RT: 7.47 min Scan# 614

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 78 Resp: 1162449

Ion Ratio Lower Upper



78 100


77 24.6 19.5 29.3



Abundance Ion 77.95 (77.65 to 78.65): VDC

Ion 76.95 (76.65 to 77.65): VDC









#41
Methacrylonitrile
Concen: 51.788 ug/l
RT: 6.46 min Scan# 512
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion: 41 Resp: 325275
Ion Ratio Lower Upper
41 100
39 57.9 47.5 71.3
67 20.4 17.8 26.6
52 12.9 10.2 15.4




Abundance Ion 40.95 (40.65 to 41.65): VDC
120000 Ion 38.95 (38.65 to 39.65): VDC
100000 Ion 67.00 (66.70 to 67.70): VDC
80000 Ion 52.00 (51.70 to 52.70): VDC






#42
1,2-Dichloroethane
Concen: 50.295 ug/l
RT: 7.60 min Scan# 627
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 62 Resp: 702740
Ion Ratio Lower Upper
62 100
98 5.1 0.0 10.2



Abundance Ion 61.95 (61.65 to 62.65): VDC
300000 Ion 97.85 (97.55 to 98.55): VDC





#43

Isopropyl Acetate

Concen: 52.814 ug/l

RT: 9.25 min Scan# 795

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 43 Resp: 352068

Ion Ratio Lower Upper




43 100

61 26.4 21.0 31.6

87 0.3 0.2 0.4



Abundance Ion 42.95 (42.65 to 43.65): VDC
Ion 60.95 (60.65 to 61.65): VDC
Ion 87.00 (86.70 to 87.70): VDC



#44

Trichloroethene

Concen: 49.753 ug/l

RT: 8.55 min Scan# 724

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion:130 Resp: 478933



Ion Ratio Lower Upper


130 100

95 101.8 0.0 208.2



Abundance Ion 129.80 (129.50 to 130.50): VDC
Ion 94.90 (94.60 to 95.60): VDC





#45

1,2-Dichloropropane

Concen: 50.655 ug/l

RT: 8.96 min Scan# 765

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :


ICVVD060519

Tot Ion: 63 Resp: 301096

Ion Ratio Lower Upper

63 100

65 32.0 21.7 32.5



Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 64.95 (64.65 to 65.65): VDC



8.96

100000

50000

0

Time--> 8.80 8.90 9.00 9.10



#46

Dibromomethane

Concen: 50.146 ug/l

RT: 9.07 min Scan# 777

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion: 93 Resp: 259317

Ion Ratio Lower Upper

93 100

95 80.5 64.2 96.2

174 106.9 84.1 126.1



Abundance Ion 92.90 (92.60 to 93.60): VDC

Ion 94.90 (94.60 to 95.60): VDC

Ion 173.80 (173.50 to 174.50): V


150000


100000

50000

0

Time--> 9.00 9.10 9.20





#47

Bromodichloromethane

Concen: 51.885 ug/l

RT: 9.39 min Scan# 809

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 83 Resp: 722690

Ion Ratio Lower Upper




83 100

85 65.8 50.2 75.4

127 10.2 7.9 11.9



Abundance Ion 82.85 (82.55 to 83.55): VDC
Ion 84.95 (84.65 to 85.65): VDC
Ion 126.80 (126.50 to 127.50): VDC



#48

Methyl methacrylate

Concen: 49.305 ug/l

RT: 9.13 min Scan# 783

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion: 41 Resp: 223526

Ion Ratio Lower Upper



41 100


69 64.2 43.8 65.8

39 77.7 64.0 96.0



Abundance Ion 40.95 (40.65 to 41.65): VDC
Ion 69.00 (68.70 to 69.70): VDC
Ion 38.95 (38.65 to 39.65): VDC









#49
1,4-Dioxane
Concen: 1002.342 ug/l
RT: 9.09 min Scan# 779
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion: 88 Resp: 36064
Ion Ratio Lower Upper
88 100
43 39.0 39.3 58.9#
58 58.7 45.3 67.9




Abundance Ion 88.00 (87.70 to 88.70): VDC
Ion 43.05 (42.75 to 43.75): VDC
Ion 57.95 (57.65 to 58.65): VDC






#50
Toluene-d8
Concen: 51.663 ug/l
RT: 10.42 min Scan# 914
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 98 Resp: 1264485
Ion Ratio Lower Upper
98 100
100 65.3 52.8 79.2



Abundance Ion 98.05 (97.75 to 98.75): VDC
Ion 100.05 (99.75 to 100.75): VDC









#51
4-Methyl-2-Pentanone
Concen: 270.323 ug/l
RT: 10.31 min Scan# 903
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 43 Resp: 1173372
Ion Ratio Lower Upper
43 100
58 32.2 26.8 40.2



#52
Toluene
Concen: 52.522 ug/l
RT: 10.52 min Scan# 924
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 92 Resp: 826947
Ion Ratio Lower Upper
92 100
91 173.1 134.7 202.1









#53
 t-1,3-Dichloropropene
 Concen: 54.270 ug/l
 RT: 10.92 min Scan# 965
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion: 75 Resp: 627637
 Ion Ratio Lower Upper
 75 100
 77 31.2 25.4 38.2



#54
 cis-1,3-Dichloropropene
 Concen: 54.733 ug/l
 RT: 10.05 min Scan# 876
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 75 Resp: 644598
 Ion Ratio Lower Upper
 75 100
 77 30.6 25.6 38.4
 39 78.4 61.2 91.8









#55
 1,1,2-Trichloroethane
 Concen: 48.068 ug/l
 RT: 11.20 min Scan# 993
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion:	97	Resp:	259649
Ion	Ratio	Lower	Upper
97	100		
83	80.5	65.9	98.9
85	54.4	42.6	63.8
99	67.0	53.4	80.2



#56
 Ethyl methacrylate
 Concen: 50.780 ug/l
 RT: 11.05 min Scan# 978
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion:	69	Resp:	305702
Ion	Ratio	Lower	Upper
69	100		
41	84.9	66.3	99.5
39	64.6	48.3	72.5









#57
 1,3-Dichloropropane
 Concen: 52.493 ug/l
 RT: 11.42 min Scan# 1015
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion: 76 Resp: 442701
 Ion Ratio Lower Upper
 76 100
 78 32.8 26.4 39.6



#58
 2-Chloroethyl Vinyl ether
 Concen: 221.190 ug/l
 RT: 9.87 min Scan# 858
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 63 Resp: 682965
 Ion Ratio Lower Upper
 63 100
 106 28.3 21.3 31.9









#59
2-Hexanone
Concen: 258.874 ug/l
RT: 11.54 min Scan# 1028
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion: 43 Resp: 852024
Ion Ratio Lower Upper
43 100
58 42.1 22.8 68.3




Abundance Ion 42.95 (42.65 to 43.65): VDC
Ion 57.95 (57.65 to 58.65): VDC






#60
Dibromochloromethane
Concen: 51.799 ug/l
RT: 11.69 min Scan# 1043
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 129 Resp: 546732
Ion Ratio Lower Upper
129 100
127 76.6 36.6 109.9



Abundance Ion 128.80 (128.50 to 129.50): V
Ion 126.80 (126.50 to 127.50): V









#61
 1,2-Dibromoethane
 Concen: 52.318 ug/l
 RT: 11.82 min Scan# 1056
 Delta R.T. 0.01 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion: 107 Resp: 346127
 Ion Ratio Lower Upper
 107 100
 109 92.3 73.8 110.8



#62
 4-Bromofluorobenzene
 Concen: 49.463 ug/l
 RT: 13.56 min Scan# 1233
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 95 Resp: 570729
 Ion Ratio Lower Upper
 95 100
 174 84.7 0.0 195.8
 176 80.3 0.0 193.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 12.38 min Scan# 1113
Delta R.T. 0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion:117 Resp: 937780
Ion Ratio Lower Upper
117 100
82 51.3 45.0 67.6
119 33.5 25.7 38.5




Abundance Ion 116.95 (116.65 to 117.65): V
Ion 82.05 (81.75 to 82.75): V
Ion 118.95 (118.65 to 119.65): V






#64
Tetrachloroethene
Concen: 50.520 ug/l
RT: 11.27 min Scan# 1000
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion:164 Resp: 427871
Ion Ratio Lower Upper
164 100
166 128.5 102.6 153.8
129 93.0 70.3 105.5
131 84.9 65.0 97.6



Abundance Ion 163.85 (163.55 to 164.55): V
Ion 165.85 (165.55 to 166.55): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 51.350 ug/l
RT: 12.41 min Scan# 1116
Delta R.T. 0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion:112 Resp: 1000627
Ion Ratio Lower Upper
112 100
114 30.7 25.0 37.6



#66
1,1,1,2-Tetrachloroethane
Concen: 50.532 ug/l
RT: 12.52 min Scan# 1127
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion:131 Resp: 473336
Ion Ratio Lower Upper
131 100
133 97.5 50.1 150.4
119 60.8 30.1 90.3









#67
Ethyl Benzene
Concen: 50.629 ug/l
RT: 12.53 min Scan# 1128
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 91 Resp: 1783157
Ion Ratio Lower Upper
91 100
106 27.8 23.0 34.4



#68
m/p-Xylenes
Concen: 101.688 ug/l
RT: 12.68 min Scan# 1143
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 106 Resp: 1249263
Ion Ratio Lower Upper
106 100
91 221.8 184.1 276.1









#69
o-Xylene
Concen: 48.868 ug/l
RT: 13.06 min Scan# 1182
Delta R.T. 0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion:106 Resp: 578447
Ion Ratio Lower Upper
106 100
91 238.5 118.8 356.5




Abundance Ion 106.05 (105.75 to 106.75): V
Ion 91.00 (90.70 to 91.70): V






#70
Styrene
Concen: 53.040 ug/l
RT: 13.08 min Scan# 1184
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion:104 Resp: 1054738
Ion Ratio Lower Upper
104 100
78 53.2 43.8 65.8
103 46.1 36.6 54.8



Abundance Ion 104.05 (103.75 to 104.75): V
Ion 77.95 (77.65 to 78.65): V
Ion 103.05 (102.75 to 103.75): V





#71

Bromoform

Concen: 51.011 ug/l

RT: 13.25 min Scan# 1201

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion:173 Resp: 330022

Ion Ratio Lower Upper

173 100

175 49.3 23.2 69.5




254 9.3 5.9 8.9#



Abundance Ion 172.80 (172.50 to 173.50): V

Ion 174.80 (174.50 to 175.50): V

Ion 253.70 (253.40 to 254.40): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 14.53 min Scan# 1331

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion:152 Resp: 517115

Ion Ratio Lower Upper

152 100

115 55.6 27.1 81.3



150 158.9 0.0 315.0




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 115.00 (114.70 to 115.70): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 50.996 ug/l

RT: 13.41 min Scan# 1218

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :


ICVVD060519

Tot Ion:105 Resp: 1833652

Ion Ratio Lower Upper




105 100

120 23.9 11.2 33.6



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.05 (119.75 to 120.75): V



#74

N-ethyl acetate

Concen: 50.880 ug/l

RT: 13.28 min Scan# 1204

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Tgt Ion: 43 Resp: 472391


Ion Ratio Lower Upper

43 100

70 30.3 20.5 30.7

55 20.5 13.8 20.8

61 22.7 15.9 23.9






Abundance Ion 42.95 (42.65 to 43.65): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 60.95 (60.65 to 61.65): V





#75

1,1,2,2-Tetrachloroethane

Concen: 51.461 ug/l

RT: 13.71 min Scan# 1248

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 83 Resp: 310183

Ion Ratio Lower Upper

83 100

131 11.4 4.7 14.0




85 67.1 32.3 96.9



Abundance Ion 82.95 (82.65 to 83.65): VDC

Ion 130.90 (130.60 to 131.60): VDC

Ion 84.95 (84.65 to 85.65): VDC



#76

1,2,3-Trichloropropane

Concen: 49.967 ug/l

RT: 13.74 min Scan# 1251

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 75 Resp: 337056

Ion Ratio Lower Upper



75 100


77 34.4 16.7 50.1



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC









#77
Bromobenzene
Concen: 51.785 ug/l
RT: 13.68 min Scan# 1245
Delta R.T. 0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion:156 Resp: 517920
Ion Ratio Lower Upper
156 100
77 112.7 70.5 211.5
158 98.8 46.9 140.6




Abundance Ion 155.90 (155.60 to 156.60): V
Ion 76.95 (76.65 to 77.65): V
Ion 157.85 (157.55 to 158.55): V






#78
n-propylbenzene
Concen: 48.316 ug/l
RT: 13.78 min Scan# 1255
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 91 Resp: 2014777
Ion Ratio Lower Upper
91 100
120 20.1 10.3 30.8



Abundance Ion 91.00 (90.70 to 91.70): V
Ion 120.05 (119.75 to 120.75): V









#79
 2-Chlorotoluene
 Concen: 50.612 ug/l
 RT: 13.85 min Scan# 1262
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion: 91 Resp: 1279540
 Ion Ratio Lower Upper
 91 100
 126 34.3 16.7 50.1



#80
 1,3,5-Trimethylbenzene
 Concen: 51.880 ug/l
 RT: 13.94 min Scan# 1271
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 105 Resp: 1605419
 Ion Ratio Lower Upper
 105 100
 120 41.1 21.1 63.4





#81

trans-1,4-Dichloro-2-butene

Concen: 57.039 ug/l

RT: 13.48 min Scan# 1225

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion: 75 Resp: 98901

Ion Ratio Lower Upper

75 100

53 121.4 113.2 169.8




89 48.3 36.9 55.3



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 53.00 (52.70 to 53.70): VDC

Ion 89.00 (88.70 to 89.70): VDC



#82

4-Chlorotoluene

Concen: 52.598 ug/l

RT: 13.96 min Scan# 1273

Delta R.T. -0.00 min

Lab File: VD062617.D


Acq: 5 Jun 2019 17:41

Tgt Ion: 91 Resp: 1564768

Ion Ratio Lower Upper



91 100


126 31.8 16.8 50.4



Abundance Ion 91.00 (90.70 to 91.70): VDC

Ion 125.95 (125.65 to 126.65): V









#83
 tert-Butylbenzene
 Concen: 52.113 ug/l
 RT: 14.20 min Scan# 1298
 Delta R.T. 0.01 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519

Tot Ion:119 Resp: 1775715




Ion	Ratio	Lower	Upper
119	100		
91	68.6	40.8	122.2
134	21.5	10.3	30.8




#84
 1,2,4-Trimethylbenzene
 Concen: 50.051 ug/l
 RT: 14.24 min Scan# 1302
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion:105 Resp: 1585306

Ion	Ratio	Lower	Upper
105	100		
120	39.6	19.9	59.7









#85
 sec-Butylbenzene
 Concen: 54.575 ug/l
 RT: 14.38 min Scan# 1316
 Delta R.T. 0.01 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion:105 Resp: 1916097
 Ion Ratio Lower Upper
 105 100
 134 18.8 8.3 24.8



#86
 p-Isopropyltoluene
 Concen: 50.046 ug/l
 RT: 14.50 min Scan# 1328
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion:119 Resp: 1689286
 Ion Ratio Lower Upper
 119 100
 134 22.4 11.2 33.5
 91 27.6 13.3 39.8









#87
 1,3-Dichlorobenzene
 Concen: 51.476 ug/l
 RT: 14.47 min Scan# 1325
 Delta R.T. 0.01 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519


Tot Ion	Ratio	Lower	Upper
146	100		
111	35.6	21.6	64.6
148	62.3	31.3	93.8



#88
 1,4-Dichlorobenzene
 Concen: 53.033 ug/l
 RT: 14.55 min Scan# 1333
 Delta R.T. 0.01 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion	Ratio	Lower	Upper
146	100		
111	37.2	20.9	62.8
148	63.6	31.5	94.5









#89
n-Butylbenzene
Concen: 49.214 ug/l
RT: 14.80 min Scan# 1359
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41




Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519


Tot Ion: 91 Resp: 1544624
Ion Ratio Lower Upper
91 100
92 58.1 29.9 89.7
134 22.6 12.1 36.3



#90
Hexachloroethane
Concen: 52.214 ug/l
RT: 15.02 min Scan# 1381
Delta R.T. 0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion: 117 Resp: 463865
Ion Ratio Lower Upper
117 100
201 91.2 34.4 103.3









#91
 1,2-Dichlorobenzene
 Concen: 47.141 ug/l
 RT: 14.81 min Scan# 1360
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD060519

Tot Ion:146 Resp: 730653
 Ion Ratio Lower Upper
 146 100
 111 45.3 21.3 64.0
 148 63.5 31.4 94.3




Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V






#92
 1,2-Dibromo-3-Chloropropane
 Concen: 50.944 ug/l
 RT: 15.38 min Scan# 1418
 Delta R.T. -0.00 min
 Lab File: VD062617.D
 Acq: 5 Jun 2019 17:41

Tgt Ion: 75 Resp: 62813
 Ion Ratio Lower Upper
 75 100
 155 65.8 41.6 124.8
 157 92.4 53.8 161.2



Abundance Ion 75.00 (74.70 to 75.70): V
 Ion 154.90 (154.60 to 155.60): V
 Ion 156.85 (156.55 to 157.55): V









#93
1,2,4-Trichlorobenzene
Concen: 50.917 ug/l
RT: 15.95 min Scan# 1476
Delta R.T. 0.01 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Instrument :
MSVOA_D
ClientSampleId :
ICVVD060519

Tot Ion:180 Resp: 567042
Ion Ratio Lower Upper
180 100
182 96.7 49.3 147.8
145 25.4 15.7 47.1




Abundance Ion 179.85 (179.55 to 180.55): V
Ion 181.85 (181.55 to 182.55): V
Ion 144.95 (144.65 to 145.65): V






#94
Hexachlorobutadiene
Concen: 53.243 ug/l
RT: 16.04 min Scan# 1485
Delta R.T. -0.00 min
Lab File: VD062617.D
Acq: 5 Jun 2019 17:41

Tgt Ion:225 Resp: 444762
Ion Ratio Lower Upper
225 100
223 62.8 30.3 91.0
227 63.3 32.4 97.0



Abundance Ion 224.75 (224.45 to 225.45): V
Ion 222.75 (222.45 to 223.45): V
Ion 226.75 (226.45 to 227.45): V





#95

Naphthalene

Concen: 53.863 ug/l

RT: 16.13 min Scan# 1494

Delta R.T. -0.00 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41

Instrument :

MSVOA_D

ClientSampleId :

ICVVD060519


Tot Ion:128 Resp: 858976

Ion Ratio Lower Upper

128 100

127 12.4 9.4 14.2




129 9.5 8.0 12.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 49.658 ug/l

RT: 16.28 min Scan# 1509

Delta R.T. 0.01 min

Lab File: VD062617.D

Acq: 5 Jun 2019 17:41


Tgt Ion:180 Resp: 422881

Ion Ratio Lower Upper

180 100

182 95.8 49.1 147.3

145 26.9 16.6 49.7



Abundance Ion 179.85 (179.55 to 180.55): V

Ion 181.85 (181.55 to 182.55): V

Ion 144.95 (144.65 to 145.65): V

