

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
 Data File : VD073476.D
 Acq On : 06 Jun 2022 15:52
 Operator : VA/SY
 Sample : VD0606SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0606SBL01

Quant Time: Jun 07 01:12:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	299346	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	513934	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	567870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	277114	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	156566	47.327	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.660%
35) Dibromofluoromethane	7.903	113	172878	50.977	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.960%
50) Toluene-d8	10.332	98	635715	50.853	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.700%
62) 4-Bromofluorobenzene	12.620	95	260029	59.271	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
Data File : VD073476.D
Acq On : 06 Jun 2022 15:52
Operator : VA/SY
Sample : VD0606SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0606SBL01

Quant Time: Jun 07 01:12:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

