

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
 Data File : VD073479.D
 Acq On : 06 Jun 2022 17:18
 Operator : VA/SY
 Sample : N3205-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-02-060322

Quant Time: Jun 07 01:13:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	281679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	505423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	538151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	259823	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	147324	47.326	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.660%
35) Dibromofluoromethane	7.903	113	165722	49.690	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.380%
50) Toluene-d8	10.332	98	610751	49.679	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.620	95	243501	56.438	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
Data File : VD073479.D
Acq On : 06 Jun 2022 17:18
Operator : VA/SY
Sample : N3205-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-02-060322

Quant Time: Jun 07 01:13:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

