

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
 Data File : VD073486.D
 Acq On : 06 Jun 2022 20:45
 Operator : VA/SY
 Sample : N3201-03
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 21500

Quant Time: Jun 07 01:14:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	239520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	448779	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	498629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	264547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	129285	48.841	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.680%
35) Dibromofluoromethane	7.909	113	149981	50.646	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.300%
50) Toluene-d8	10.332	98	556367	50.967	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.940%
62) 4-Bromofluorobenzene	12.620	95	252129	65.814	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	131.620%
Target Compounds						
					Qvalue	
16) Acetone	4.173	43	7737	16.875	ug/l	89
73) Isopropylbenzene	12.473	105	53853	2.856	ug/l #	68
86) p-Isopropyltoluene	13.496	119	20482	1.222	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
Data File : VD073486.D
Acq On : 06 Jun 2022 20:45
Operator : VA/SY
Sample : N3201-03
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
21500

Quant Time: Jun 07 01:14:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

