

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
 Data File : VD073487.D
 Acq On : 06 Jun 2022 21:14
 Operator : VA/SY
 Sample : N3225-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-01-06062022

Quant Time: Jun 07 01:14:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	96635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	173956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	233254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	138508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	74690	69.937	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.880%	
35) Dibromofluoromethane	7.908	113	66273	57.735	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.480%	
50) Toluene-d8	10.332	98	214988	50.808	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.620%	
62) 4-Bromofluorobenzene	12.620	95	119110	80.211	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 160.420%#	
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	2720	14.704	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060622\
Data File : VD073487.D
Acq On : 06 Jun 2022 21:14
Operator : VA/SY
Sample : N3225-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-01-06062022

Quant Time: Jun 07 01:14:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

