

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060623\
 Data File : VD076336.D
 Acq On : 06 Jun 2023 13:01
 Operator : KP/SY
 Sample : 03032-04
 Misc : 4.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-21-0A

Quant Time: Jun 07 06:10:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	157800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	294054	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	263408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	102956	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	100552	71.118	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.240%
35) Dibromofluoromethane	7.805	113	135208	72.418	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	144.840%
50) Toluene-d8	10.269	98	403828	54.324	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.640%
62) 4-Bromofluorobenzene	12.575	95	108371	55.088	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060623\
Data File : VD076336.D
Acq On : 06 Jun 2023 13:01
Operator : KP/SY
Sample : 03032-04
Misc : 4.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-21-0A

Quant Time: Jun 07 06:10:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

