

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060623\
 Data File : VD076346.D
 Acq On : 06 Jun 2023 18:03
 Operator : KP/SY
 Sample : 03042-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Jun 07 06:11:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	1968	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.763	114	3603	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.587	117	2905	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.528	152	1157	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	2428	137.695	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	275.400%#
35) Dibromofluoromethane	7.793	113	1322	57.788	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.580%
50) Toluene-d8	10.263	98	3562	39.107	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	78.220%
62) 4-Bromofluorobenzene	12.575	95	932	38.665	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060623\
Data File : VD076346.D
Acq On : 06 Jun 2023 18:03
Operator : KP/SY
Sample : 03042-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Jun 07 06:11:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

