

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060718\
 Data File : VD057970.D
 Acq On : 7 Jun 2018 14:29
 Operator : JC/SY
 Sample : J3320-01
 Misc : 6.24g/5ml/MSVOA D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S39-0177

Quant Time: Jun 08 07:29:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	817606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	1022246	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.68	117	748482	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	264962	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.01	65	300833	36.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.60%	
35) Dibromofluoromethane	5.55	113	344672	43.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.16%	
50) Toluene-d8	8.70	98	816011	43.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.80%	
62) 4-Bromofluorobenzene	11.93	95	274678	33.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.98%	
Target Compounds						
16) Acetone	2.63	43	93951	79.778	ug/l	96
21) trans-1,2-Dichloroethene	3.39	96	14616	2.001	ug/l #	73
25) 2-Butanone	4.84	43	43043	16.083	ug/l #	78
27) cis-1,2-Dichloroethene	4.81	96	97333	11.236	ug/l	91
37) Ethyl Acetate	4.92	43	60223	9.800	ug/l #	89
44) Trichloroethene	6.99	130	10768	1.401	ug/l	86
84) 1,2,4-Trimethylbenzene	12.62	105	22116	1.503	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD060718\
Data File : VD057970.D
Acq On : 7 Jun 2018 14:29
Operator : JC/SY
Sample : J3320-01
Misc : 6.24g/5ml/MSVOA D/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S39-0177

Quant Time: Jun 08 07:29:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration

