

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060718\
 Data File : VD057973.D
 Acq On : 7 Jun 2018 15:54
 Operator : JC/SY
 Sample : J3320-04
 Misc : 6.26g/5ml/MSVOA D/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S39-0179

Quant Time: Jun 08 07:36:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	979297	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.69	114	1266834	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.68	117	985042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	442593	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.01	65	365398	37.32	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	74.64%	
35) Dibromofluoromethane	5.54	113	384011	39.18	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	78.36%	
50) Toluene-d8	8.70	98	875901	37.59	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	75.18%	
62) 4-Bromofluorobenzene	11.92	95	356662	35.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.18%	
Target Compounds						
16) Acetone	2.61	43	55945	39.662	ug/l	96
17) Carbon Disulfide	2.76	76	16367	1.150	ug/l #	91
25) 2-Butanone	4.83	43	29847	9.311	ug/l #	86
27) cis-1,2-Dichloroethene	4.80	96	37712	3.635	ug/l	81
37) Ethyl Acetate	4.91	43	24260	3.186	ug/l #	92
52) Toluene	8.79	92	67959	3.632	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD060718\
Data File : VD057973.D
Acq On : 7 Jun 2018 15:54
Operator : JC/SY
Sample : J3320-04
Misc : 6.26g/5ml/MSVOA D/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S39-0179

Quant Time: Jun 08 07:36:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration

